

Your P.O. #: BN500636
 Your Project #: BISHOP SCALABRINI
 Site:500019397
 Your C.O.C. #: na

Attention: Roland Lawton
 Dufferin Peel Catholic District...
 School Board
 40 Matheson Blvd W
 Mississauga, ON
 L5R 1C5

Report Date: 2008/09/23

CERTIFICATE OF ANALYSIS

MAXXAM JOB #: A8A4220
Received: 2008/09/12, 12:31

Sample Matrix: Water
 # Samples Received: 2

Analyses	Quantity	Date Extracted	Date Analyzed	Laboratory Method	Method Reference
Metals Analysis by ICPMS (as received) 0	2	2008/09/22	2008/09/22	CAM SOP-00447	EPA 6020

* RPDs calculated using raw data. The rounding of final results may result in the apparent difference.

(1) Metals analysis was performed on the sample 'as received'.

Encryption Key

Please direct all questions regarding this Certificate of Analysis to your Project Manager.

JAMES ASPIN, Project Manager
 Email: James.Aspin@maxxamAnalytics.com
 Phone# (905) 817-5700 Ext:5771

=====

Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.

For Service Group specific validation please refer to the Validation Signature Page

Total cover pages: 1

Maxxam Job #: A8A4220
 Report Date: 2008/09/23

Dufferin Peel Catholic District...
 Client Project #: BISHOP SCALABRINI
 Project name: 500019397
 Your P.O. #: BN500636
 Sampler Initials: CT

ELEMENTS BY ATOMIC SPECTROSCOPY (WATER)

Maxxam ID			AL9629		AL9630		
Sampling Date			2008/09/12 06:30		2008/09/12 07:05		
COC Number			na		na		
	Units	Criteria	BISHOP SCALABRINI KITCHEN- STAFFROOM STANDING	QC Batch	BISHOP SCALABRINI KITCHEN- STAFFROOM FLUSHED	RDL	QC Batch

Metals							
. Lead (Pb)	mg/L	0.01	0.0073	1620450	0.0023	0.0005	1620430

RDL = Reportable Detection Limit
 QC Batch = Quality Control Batch
 CRITERIA: Ontario Regulation 243/07 made under the Safe Drinking Water Act, 2002.
 Released June 7, 2007

Maxxam Job #: A8A4220
Report Date: 2008/09/23

Dufferin Peel Catholic District...
Client Project #: BISHOP SCALABRINI
Project name: 500019397
Your P.O. #: BN500636
Sampler Initials: CT

Package 1	12.0°C
-----------	--------

Each temperature is the average of up to three cooler temperatures taken at receipt

GENERAL COMMENTS

Results relate only to the items tested.

Dufferin Peel Catholic District...
 Attention: Roland Lawton
 Client Project #: BISHOP SCALABRINI
 P.O. #: BN500636
 Project name: 500019397

Quality Assurance Report
 Maxxam Job Number: MA8A4220

QA/QC Batch	QC Type	Parameter	Date Analyzed yyyy/mm/dd	Value	Recovery	Units	QC Limits
1620430 GBU	MATRIX SPIKE	. Lead (Pb)	2008/09/22		99	%	80 - 120
	Spiked Blank	. Lead (Pb)	2008/09/22		95	%	85 - 115
	Method Blank	. Lead (Pb)	2008/09/22	ND, RDL=0.0005		mg/L	
	RPD	. Lead (Pb)	2008/09/22	NC		%	25
1620450 GBU	MATRIX SPIKE	. Lead (Pb)	2008/09/22		100	%	80 - 120
	Spiked Blank	. Lead (Pb)	2008/09/22		95	%	85 - 115
	Method Blank	. Lead (Pb)	2008/09/22	ND, RDL=0.0005		mg/L	

ND = Not detected
 NC = Non-calculable
 RPD = Relative Percent Difference
 SPIKE = Fortified sample

Validation Signature Page

Maxxam Job #: A8A4220

The analytical data and all QC contained in this report were reviewed and validated by the following individual(s).

 

EWA PRANJIC, M.Sc., C.Chem, Scientific Specialist

=====

Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.



6740 Campobello Rd
Mississauga
Ontario L5N 2L8

12-Sep-08 12:31

JAMES ASPIN



A8A4220

NNA ENV-987

DRINKING WATER SUBMISSION

CUSTODY RECORD

Samples are Regulated
Under Reg 243

Maxxam Analytics Inc.
CAM FCD-00196/2

Report to:

Company Name: Dufferin-Peel Catholic District School Board
Client Contact: Roland Lawton, General Manager of Physical Plant Facilities
Address: 5685 Keaton Cres., Mississauga ON L5R 3H5
E-Mail: roland.lawton@dpcdsb.org Tel: 905-890-0708 ext. 23204
Project: Regulation 243 Fax: 905-890-5841
Quote #: A83456 P.O.#: BN500636
Lab Service Rep.: James Aspin

Invoice to: Same as "Report to"

PAGE ____ OF ____

ANALYSIS REQUESTED

Specify School Type

School Private School Day nursery

M.O.E. Assigned Water System Number

500019397

08 SEP 12 12:31

Sample Identification	Sample Type (Please Specify Standing / Flushed)	Date Sampled	Time Sampled	Sampler Initials	Field Chlorine		No. of Bottles	Lead	School	Private School	Day nursery	M.O.E. Assigned Water System Number
					Total	Free						
Bishop Scalabrini												
Kitchen - Staffroom	Standing	12-9-08	6:30AM	C.J.			1	X	X			
Kitchen - Staffroom	Flushed	12-9-08	7:05AM	C.J.			1	X	X			

DW METALS PRESERVATION CHECKED

notification of adverse results must be reported to the SAC, MOH & Waterworks.
OPERATIVE that the areas shaded in grey be filled out completely before submitting to Maxxam Analytics.

TAT (Turnaround Time)	Notification Information for School, Private Schools & Day Nurseries		LABORATORY USE ONLY
RUSH TAT MUST HAVE PRIOR APPROVAL	Facility Name: Bishop Scalabrini	Medical Officer of Health	Received By: <i>Naglaa Mikhail</i>
<input checked="" type="checkbox"/> Normal (5-15 Working Days)	Facility Address: 225 Central Parkway W., Mississauga, ON L5B 3J5	Region: Peel Region	Date: 08/09/12 Time: 12:31
<input type="checkbox"/> RUSH (Specify Below)		Tel #:	Comments: 12/12/12 insufficient cooling
Date Due:	Tel: 905-896-3665 Fax: 905-896-8858	Fax #:	
PN #	E-mail:	Email:	
Relinquished by (Sign Name): <i>Carlos Tome</i>	Date: <i>SEPT. 12.08</i>	Laboratory Prepared Container:	<input type="checkbox"/> Yes <input type="checkbox"/> No
Relinquished by (Print Name): <i>CARLOS TOME</i>	Time: <i>7:10 AM</i>		

NOTE: Samples submitted to Maxxam Analytics must be inspected by a SIF (Sample Inspection) before use.
It is the Sampler's responsibility to ensure adequate cooling is provided so that the samples are preserved.
Please ensure that a separate COC document is used for each sample.
Resolved By: _____
Date: _____