

Your P.O. #: BN500636
Your Project #: BLESSED TRINITY
Site:500019462
Your C.O.C. #: NA

Attention: Roland Lawton
Dufferin Peel Catholic District...
School Board
40 Matheson Blvd W
Mississauga, ON
L5R 1C5

Report Date: 2008/09/12

CERTIFICATE OF ANALYSIS

MAXXAM JOB #: A8A1898
Received: 2008/09/09, 08:02

Sample Matrix: Water
Samples Received: 2

Analyses	Quantity	Date Extracted	Date Analyzed	Laboratory Method	Method Reference
Metals Analysis by ICPMS (as received) 0	2	2008/09/11	2008/09/11	CAM SOP-00447	EPA 6020

* RPDs calculated using raw data. The rounding of final results may result in the apparent difference.

(1) Metals analysis was performed on the sample 'as received'.

Encryption Key

Please direct all questions regarding this Certificate of Analysis to your Project Manager.

JAMES ASPIN, Project Manager
Email: James.Aspin@maxxamAnalytics.com
Phone# (905) 817-5700 Ext:5771

=====

Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.

For Service Group specific validation please refer to the Validation Signature Page

Total cover pages: 1

Maxxam Job #: A8A1898
 Report Date: 2008/09/12

Dufferin Peel Catholic District...
 Client Project #: BLESSED TRINITY
 Project name: 500019462
 Your P.O. #: BN500636

ELEMENTS BY ATOMIC SPECTROSCOPY (WATER)

Maxxam ID			AK8253	AK8254		
Sampling Date			2008/09/09 06:40	2008/09/09 07:15		
COC Number			NA	NA		
	Units	Criteria	BLESSED TRINITY CATHOLIC STANDING	BLESSED TRINITY CATHOLIC FLUSHED	RDL	QC Batch

Metals						
. Lead (Pb)	mg/L	0.01	0.0017	0.0008	0.0005	1610625

RDL = Reportable Detection Limit
 QC Batch = Quality Control Batch
 CRITERIA: Ontario Regulation 243/07 made under the Safe Drinking Water Act, 2002.
 Released June 7, 2007

Maxxam Job #: A8A1898
Report Date: 2008/09/12

Dufferin Peel Catholic District...
Client Project #: BLESSED TRINITY
Project name: 500019462
Your P.O. #: BN500636

Package 1	20.3°C
-----------	--------

Each temperature is the average of up to three cooler temperatures taken at receipt

GENERAL COMMENTS

Results relate only to the items tested.

Dufferin Peel Catholic District...
 Attention: Roland Lawton
 Client Project #: BLESSED TRINITY
 P.O. #: BN500636
 Project name: 500019462

Quality Assurance Report
 Maxxam Job Number: MA8A1898

QA/QC Batch	QC Type	Parameter	Date Analyzed yyyy/mm/dd	Value	Recovery	Units	QC Limits
1610625 HRE	MATRIX SPIKE	. Lead (Pb)	2008/09/11		96	%	80 - 120
	Spiked Blank	. Lead (Pb)	2008/09/11		97	%	85 - 115
	Method Blank	. Lead (Pb)	2008/09/11	ND, RDL=0.0005		mg/L	
	RPD	. Lead (Pb)	2008/09/11	NC		%	25

ND = Not detected
 NC = Non-calculable
 RPD = Relative Percent Difference
 SPIKE = Fortified sample

Validation Signature Page

Maxxam Job #: A8A1898

The analytical data and all QC contained in this report were reviewed and validated by the following individual(s).



BRAD NEWMAN, Scientific Specialist

=====

Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.



6740 Campobello Rd
Mississauga
Ontario L5N 2L8

REGULATION 243 DRINKING WATER SUBM CHAIN OF CUSTODY RECORD

Tel: 905-817-5700
Fax: 905-817-5779
Toll Free: 1-800-563-6266

Samples are Regulated
Under Reg 243

9-Sep-08 08:02
JAMES ASPIN
A8A1898
M_O ENV-951

Report to:

Company Name: Dufferin-Peel Catholic District School Board
Client Contact: Roland Lawton, General Manager of Physical Plant Facilities
Address: 5685 Keaton Cres., Mississauga ON L5R 3H5
E-Mail: roland.lawton@dpcdsb.org Tel: 905-890-0708 ext. 23204
Project: Regulation 243 Fax: 905-890-5841
Quote #: A83456 P.O.#: BN500636
Lab Service Rep.: James Aspin

Invoice to: Same as "Report to"

PAGE		OF
ANALYSIS REQUESTED		
Specify School Type	M.O.E. Assigned Water System Number	
School	Private School	Day nursery
		500019462

Sample Identification	Sample Type (Please Specify Standing / Flushed)	Date Sampled	Time Sampled	Sampler Initials	Field Chlorine		No. of Bottles	Lead
					Total	Free		
Blessed Trinity Catholic Kitchen - Staffroom	Standing	Sept 9	6:40 AM				1	X
Kitchen - Staffroom	Flushed	Sept 9					1	X

SIF : Sample Inspection
Resolved By: _____

* In accordance with the Safe Drinking Water Act (2002), notification of adverse results must be provided to the appropriate authority. Failure to do so is an offence under the act. It is IMPERATIVE that the areas shaded in grey be completed before submitting to Maxxam Analytics.

TAT (Turnaround Time)	Notification Information for School, Private Schools & Day Nurseries		LABORATORY USE ONLY	
RUSH TAT MUST HAVE PRIOR APPROVAL	Facility Name: Blessed Trinity Catholic	Medical Officer of Health	Received By: <i>[Signature]</i>	
<input checked="" type="checkbox"/> Normal (5-15 Working Days)	Facility Address: 2495 Credit Valley Road	Region: Peel Region	Date: 08/09/09	Time: 8:02
<input type="checkbox"/> RUSH (Specify Below)	Mississauga, ON L5M 4G8	Tel #:	Comments: 20/20/21c no ice pads.	
Date Due:	Tel: 905-828-4167 Fax: 905-828-5847	Fax #:		
PN #	E-mail:	Email:	Laboratory Prepared Container: <input type="checkbox"/> Yes <input type="checkbox"/> No	
Relinquished by (Sign Name): <i>[Signature]</i>	Date: Sept 9-09	Time: 8:05		
Relinquished by (Print Name): PAOLO TENACCI				

NOTE: Samples submitted to Maxxam Analytics must be received by 12:00 PM. It is the Sampler's responsibility to ensure adequate cooling is provided. Please ensure that a separate COC document is used for

DW METALS
PRESERVATION
CHECKED

08 SEP 9 8:02