

Your P.O. #: BN500116  
Your Project #: CORPUS CHRISTI  
Site:500018968  
Your C.O.C. #: na

**Attention: Roland Lawton**  
Dufferin Peel Catholic District...  
School Board  
40 Matheson Blvd W  
Mississauga, ON  
L5R 1C5

**Report Date: 2008/09/03**

**CERTIFICATE OF ANALYSIS**

**MAXXAM JOB #: A896605**  
**Received: 2008/08/28, 11:07**

Sample Matrix: Water  
# Samples Received: 2

Analyses	Quantity	Date Extracted	Date Analyzed	Laboratory Method	Method Reference
Metals Analysis by ICPMS (as received) ¶	2	2008/09/02	2008/09/03	CAM SOP-00447	EPA 6020

\* RPDs calculated using raw data. The rounding of final results may result in the apparent difference.

(1) Metals analysis was performed on the sample 'as received'.

**Encryption Key**

Please direct all questions regarding this Certificate of Analysis to your Project Manager.

JAMES ASPIN, Project Manager  
Email: James.Aspin@maxxamAnalytics.com  
Phone# (905) 817-5700 Ext:5771

=====  
Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.

For Service Group specific validation please refer to the Validation Signature Page

Total cover pages: 1

Maxxam Job #: A896605  
Report Date: 2008/09/03

Dufferin Peel Catholic District...  
Client Project #: CORPUS CHRISTI  
Project name: 500018968  
Your P.O. #: BN500116

**ELEMENTS BY ATOMIC SPECTROSCOPY (WATER)**

Maxxam ID			AI0308		
Sampling Date			2008/08/19 07:10		
COC Number			na		
	<b>Units</b>	<b>Criteria</b>	<b>CORPUS CHRISTI-KITCHEN-STAFFROOM-STA NDING</b>	<b>RDL</b>	<b>QC Batch</b>

<b>Metals</b>					
. Lead (Pb)	mg/L	<b>0.01</b>	0.0078	0.0005	1601891
<p>RDL = Reportable Detection Limit            QC Batch = Quality Control Batch            CRITERIA: Ontario Regulation 243/07 made under the Safe Drinking Water Act, 2002.            Released June 7, 2007</p>					

Maxxam ID			AI0309		
Sampling Date			2008/08/19 07:45		
COC Number			na		
	<b>Units</b>	<b>Criteria</b>	<b>CORPUS CHRISTI-KITCHEN-STAFFROOM-FLU SHED</b>	<b>RDL</b>	<b>QC Batch</b>

<b>Metals</b>					
. Lead (Pb)	mg/L	<b>0.01</b>	0.0046	0.0005	1601891
<p>RDL = Reportable Detection Limit            QC Batch = Quality Control Batch            CRITERIA: Ontario Regulation 243/07 made under the Safe Drinking Water Act, 2002.            Released June 7, 2007</p>					

Maxxam Job #: A896605  
Report Date: 2008/09/03

Dufferin Peel Catholic District...  
Client Project #: CORPUS CHRISTI  
Project name: 500018968  
Your P.O. #: BN500116

Package 1	5.0°C
-----------	-------

Each temperature is the average of up to three cooler temperatures taken at receipt

**GENERAL COMMENTS**

**Results relate only to the items tested.**

Dufferin Peel Catholic District...  
Attention: Roland Lawton  
Client Project #: CORPUS CHRISTI  
P.O. #: BN500116  
Project name: 500018968

Quality Assurance Report  
Maxxam Job Number: MA896605

QA/QC Batch	QC Type	Parameter	Date Analyzed yyyy/mm/dd	Value	Recovery	Units	QC Limits
1601891 MIL	MATRIX SPIKE	. Lead (Pb)	2008/09/03		101	%	80 - 120
	Spiked Blank	. Lead (Pb)	2008/09/03		98	%	85 - 115
	Method Blank	. Lead (Pb)	2008/09/03	ND, RDL=0.0005		mg/L	
	RPD	. Lead (Pb)	2008/09/03	NC		%	25

ND = Not detected  
NC = Non-calculable  
RPD = Relative Percent Difference  
SPIKE = Fortified sample

**Validation Signature Page**

**Maxxam Job #: A896605**

---

The analytical data and all QC contained in this report were reviewed and validated by the following individual(s).

*Cristina Nervo*

---

CHRISTINA NERVO, Scientific Services

=====

Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.



6740 Campobello Rd  
Mississauga  
Ontario L5N 2L8

# REGULATION 243 DRINKING WATER SUBM CHAIN OF CUSTODY RECORD

Tel: 905-817-5700  
Fax: 905-817-5779  
Toll Free: 1-800-563-6266

Samples are Regulated  
Under Reg 243

28-Aug-08 11:07

JAMES ASPIN



A896605

CCO ENV-719

Maxxam Analytics Inc.  
CAM FCD-00196/2

**Report to:**

Company Name: Dufferin-Peel Catholic District School Board  
Client Contact: Roland Lawton, General Manager of Physical Plant Facilities  
Address: 5685 Keaton Cres., Mississauga ON L5R 3H5  
E-Mail: roland.lawton@dpcdsb.org Tel: 905-890-0708 ext. 23204  
Project: Regulation 243 Fax: 905-890-5841  
Quote #: A83456 P.O.#: BN500636  
Lab Service Rep.: James Aspen

Invoice to:  Same as "Report to"

PAGE \_\_\_\_ OF \_\_\_\_

**ANALYSIS REQUESTED**

Specify School Type

Private School Day nursery

M.O.E. Assigned Water System Number

500018968

Sample Identification	Sample Type (Please Specify Standing / Flushed)	Date Sampled	Time Sampled	Sampler Initials	Field Chlorine		No. of Bottles	Lead
					Total	Free		
Corpus Christi Kitchen - Staffroom	Standing	19/08/08	7:10 AM	JH			1	X
Kitchen - Staffroom	Flushed	19/08/08	7:45 AM	JH			1	X

DW METALS  
PRESERVATION  
CHECKED

\* In accordance with the Safe Drinking Water Act (2002), notification of adverse results must be reported to the SAC, MOH & Waterworks. Failure to do so is an offence under the act. It is IMPERATIVE that the areas shaded in grey be filled out completely before submitting to Maxxam

TAT (Turnaround Time)	Notification Information for School, Private Schools & Day Nurseries		LABORATORY USE ONLY
<b>RUSH TAT MUST HAVE PRIOR APPROVAL</b>	Facility Name: Corpus Christi	Medical Officer of Health	Received By: <i>[Signature]</i> 11:00
<input checked="" type="checkbox"/> Normal (5-15 Working Days)	Facility Address: 4155 Elora Drive	Region: Peel Region	Date: 08/08/08 Time: 12:35
<input type="checkbox"/> RUSH (Specify Below)	Mississauga, ON L5B 3N4	Tel #: 905 791 7800	Comments: 5/5/5
Date Due:	Tel: 905-897-7037 Fax: 905-566-4333	Fax #: 905 789 0328	
PN #	E-mail:	Email:	
Relinquished by (Sign Name):	Date: 19 06 AUG 08	Laboratory Prepared Container:	<input checked="" type="checkbox"/> Yes <input type="checkbox"/> No
Relinquished by (Print Name):	Time: 7:50 AM		

NOTE: Samples submitted to Maxxam Analytics must be received below 10°C.

It is the Sampler's responsibility to ensure adequate cooling is provided. Please ensure that a separate COC document is used

SIF : Sample Inspection

Page 6 of 6  
Resolved  By: \_\_\_\_\_

08 AUG 28 11:07