

Your Project #: DIVINE MERCY
Site:500018981
Your C.O.C. #: na

Attention: Roland Lawton
Dufferin Peel Catholic District...
School Board
40 Matheson Blvd W
Mississauga, ON
L5R 1C5

Report Date: 2008/07/28

CERTIFICATE OF ANALYSIS

MAXXAM JOB #: A878424
Received: 2008/07/17, 14:30

Sample Matrix: Water
Samples Received: 2

Analyses	Quantity	Date Extracted	Date Analyzed	Laboratory Method	Method Reference
Metals Analysis by ICPMS (as received) ①	1	2008/07/25	2008/07/25	CAM SOP-00447	EPA 6020
Metals Analysis by ICPMS (as received) ①	1	2008/07/25	2008/07/26	CAM SOP-00447	EPA 6020

* RPDs calculated using raw data. The rounding of final results may result in the apparent difference.

(1) Metals analysis was performed on the sample 'as received'.

Encryption Key

Please direct all questions regarding this Certificate of Analysis to your Project Manager.

JAMES ASPIN, Project Manager
Email: James.Aspin@maxxamAnalytics.com
Phone# (905) 817-5700 Ext:5771

=====
Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.

For Service Group specific validation please refer to the Validation Signature Page

Total cover pages: 1

Maxxam Job #: A878424
Report Date: 2008/07/28

Dufferin Peel Catholic District...
Client Project #: DIVINE MERCY
Project name: 500018981
Sampler Initials: KW

ELEMENTS BY ATOMIC SPECTROSCOPY (WATER)

Maxxam ID			Z97155		Z97156		
Sampling Date			2008/07/17 07:30		2008/07/17 08:05		
COC Number			na		na		
	Units	Criteria	DIVINE MERCY KITCHEN- STAFFROOM STANDING	QC Batch	DIVINE MERCY KITCHEN- STAFFROOM FLUSHED	RDL	QC Batch

Metals							
. Lead (Pb)	mg/L	0.01	0.0018	1570417	0.0006	0.0005	1571011

RDL = Reportable Detection Limit
QC Batch = Quality Control Batch
CRITERIA: Ontario Regulation 243/07 made under the Safe Drinking Water Act, 2002.
Released June 7, 2007

Maxxam Job #: A878424
Report Date: 2008/07/28

Dufferin Peel Catholic District...
Client Project #: DIVINE MERCY
Project name: 500018981
Sampler Initials: KW

Package 1	9.3°C
-----------	-------

Each temperature is the average of up to three cooler temperatures taken at receipt

GENERAL COMMENTS

Results relate only to the items tested.

Dufferin Peel Catholic District...
Attention: Roland Lawton
Client Project #: DIVINE MERCY
P.O. #:
Project name: 500018981

Quality Assurance Report
Maxxam Job Number: MA878424

QA/QC Batch	QC Type	Parameter	Date Analyzed yyyy/mm/dd	Value	Recovery	Units	QC Limits
1570417 HRE	MATRIX SPIKE	. Lead (Pb)	2008/07/25		100	%	80 - 120
	Spiked Blank	. Lead (Pb)	2008/07/25		97	%	85 - 115
	Method Blank	. Lead (Pb)	2008/07/25	ND, RDL=0.5		ug/L	
	RPD	. Lead (Pb)	2008/07/25	1.2		%	25
1571011 HRE	MATRIX SPIKE	. Lead (Pb)	2008/07/26		100	%	80 - 120
	Spiked Blank	. Lead (Pb)	2008/07/26		100	%	85 - 115
	Method Blank	. Lead (Pb)	2008/07/26	ND, RDL=0.5		ug/L	
	RPD	. Lead (Pb)	2008/07/26	1.5		%	25

ND = Not detected
RPD = Relative Percent Difference
SPIKE = Fortified sample

Validation Signature Page

Maxxam Job #: A878424

The analytical data and all QC contained in this report were reviewed and validated by the following individual(s).

Cristina Nervo

CHRISTINA NERVO, Scientific Services

=====

Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.

SIF : Sample Inspection

Resolved By: _____

Date: _____

6740 Campobello Rd
Mississauga
Ontario L5N 2L8

17-Jul-08 14:30

JAMES ASPIN



A878424

NNA

ENV-998

Toll Free: 1-800-563-6266

DRINKING WATER
CUSTODY REC

Samples are Regulated
Under Reg 243

DW METALS
PRESERVATION
CHECKED

Maxxam Analytics Inc.
CAM FCD-00196/2

Report to:

Company Name:	Dufferin-Peel Catholic District School Board		
Client Contact:	Roland Lawton, General Manager of Physical Plant Facilities		
Address:	5685 Keaton Cres., Mississauga ON L5R 3H5		
E-Mail:	<u>roland.lawton@dpcdsb.org</u>	Tel:	905-890-0708 ext. 23204
Project:	Regulation 243	Fax:	905-890-5841
Quote #:	A83456	P.O.#:	BN500636
Lab Service Rep.:	James Aspen		

Invoice to: Same as "Report to"

PAGE ____ OF ____

ANALYSIS REQUESTED			
Specify School Type			
School	Private School	Day nursery	M.O.E. Assigned Water System Number
			500018981

Sample Identification	Sample Type (Please Specify Standing / Flushed)	Date Sampled	Time Sampled	Sampler Initials	Field Chlorine		No. of Bottles	Lead
					Total	Free		
Divine Mercy								
Kitchen - Staffroom	Standing	Jul 17	7:30				1	x
Kitchen - Staffroom	Flushed	Jul 17	8:05				1	x

* In accordance with the Safe Drinking Water Act (2002), notification of adverse results must be reported to the SAC, MOH & Waterworks. Failure to do so is an offence under the act. It is **IMPERATIVE** that the areas shaded in grey be filled out completely before submitting to Maxxam Analytics.

TAT (Turnaround Time)	Notification Information for School, Private Schools & Day Nurseries		LABORATORY USE ONLY
RUSH TAT MUST HAVE PRIOR APPROVAL	Facility Name: Divine Mercy	Medical Officer of Health	Received By: <i>M. Spangola</i>
<input checked="" type="checkbox"/> Normal (5-15 Working Days)	Facility Address: 2840 Duncairn Drive	Region: Peel Region	Date: 08/07/17 Time: 14:30
<input type="checkbox"/> RUSH (Specify Below)	Mississauga, ON L5M 5C6	Tel #:	Comments: 9/9/10' 08 JUL 17 14:30
Date Due:	Tel: 905-812-5445 Fax: 905-812-4922	Fax #:	
PN #	E-mail:	Email:	
Relinquished by (Sign Name): <i>K Williams</i>	Date: <i>July 17/08</i>	Laboratory Prepared Container:	<input type="checkbox"/> Yes <input type="checkbox"/> No
Relinquished by (Print Name): <i>Kim Williams</i>	Time: <i>2:47 pm</i>		

NOTE: Samples submitted to Maxxam Analytics must be received below 10°C. It is the Sampler's responsibility to ensure adequate cooling is provided so that this receiving temperature is met. Please ensure that a separate COC document is used for each individual waterworks.