

Your P.O. #: BN500116  
Your Project #: FATHER MICHAEL J MCGIVREY  
Site:500012260  
Your C.O.C. #: na

**Attention: Roland Lawton**  
Dufferin Peel Catholic District...  
School Board  
40 Matheson Blvd W  
Mississauga, ON  
L5R 1C5

**Report Date: 2008/07/25**

**CERTIFICATE OF ANALYSIS**

**MAXXAM JOB #: A878374**  
**Received: 2008/07/17, 14:04**

Sample Matrix: Water  
# Samples Received: 2

Analyses	Quantity	Date Extracted	Date Analyzed	Laboratory Method	Method Reference
Metals Analysis by ICPMS (as received) 0	2	2008/07/24	2008/07/24	CAM SOP-00447	EPA 6020

\* RPDs calculated using raw data. The rounding of final results may result in the apparent difference.

(1) Metals analysis was performed on the sample 'as received'.

**Encryption Key**

Please direct all questions regarding this Certificate of Analysis to your Project Manager.

JAMES ASPIN, Project Manager  
Email: James.Aspin@maxxamAnalytics.com  
Phone# (905) 817-5700 Ext:5771

=====  
Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.

For Service Group specific validation please refer to the Validation Signature Page

Total cover pages: 1

Maxxam Job #: A878374  
Report Date: 2008/07/25

Dufferin Peel Catholic District...  
Client Project #: FATHER MICHAEL J MCGIVREY  
Project name: 500012260  
Your P.O. #: BN500116

**ELEMENTS BY ATOMIC SPECTROSCOPY (WATER)**

Maxxam ID			Z96839		
Sampling Date			2008/07/17 06:30		
COC Number			na		
	<b>Units</b>	<b>Criteria</b>	<b>FATHER MICHAEL J. MCGIVNEY-KITCHEN-STAFFROOM-S TANDING</b>	<b>RDL</b>	<b>QC Batch</b>

<b>Metals</b>					
. Lead (Pb)	mg/L	<b>0.01</b>	0.0018	0.0005	1569380
<p>RDL = Reportable Detection Limit            QC Batch = Quality Control Batch            CRITERIA: Ontario Regulation 243/07 made under the Safe Drinking Water Act, 2002.            Released June 7, 2007</p>					

Maxxam ID			Z96840		
Sampling Date			2008/07/17 07:07		
COC Number			na		
	<b>Units</b>	<b>Criteria</b>	<b>FATHER MICHAEL J. MCGIVNEY-KITCHEN-STAFFROOM-F LUSHED</b>	<b>RDL</b>	<b>QC Batch</b>

<b>Metals</b>					
. Lead (Pb)	mg/L	<b>0.01</b>	0.0008	0.0005	1569380
<p>RDL = Reportable Detection Limit            QC Batch = Quality Control Batch            CRITERIA: Ontario Regulation 243/07 made under the Safe Drinking Water Act, 2002.            Released June 7, 2007</p>					

Maxxam Job #: A878374  
Report Date: 2008/07/25

Dufferin Peel Catholic District...  
Client Project #: FATHER MICHAEL J MCGIVREY  
Project name: 500012260  
Your P.O. #: BN500116

Package 1	10.0°C
-----------	--------

Each temperature is the average of up to three cooler temperatures taken at receipt

**GENERAL COMMENTS**

**Results relate only to the items tested.**

Dufferin Peel Catholic District...  
Attention: Roland Lawton  
Client Project #: FATHER MICHAEL J MCGIVREY  
P.O. #: BN500116  
Project name: 500012260

**Quality Assurance Report**  
Maxxam Job Number: MA878374

QA/QC Batch			Date Analyzed					
Num Init	QC Type	Parameter	yyyy/mm/dd	Value	Recovery	Units	QC Limits	
1569380 HRE	MATRIX SPIKE	. Lead (Pb)	2008/07/24		99	%	80 - 120	
	Spiked Blank	. Lead (Pb)	2008/07/24		98	%	85 - 115	
	Method Blank	. Lead (Pb)	2008/07/24	ND, RDL=0.5		ug/L		
	RPD	. Lead (Pb)	2008/07/24	NC		%	25	

ND = Not detected  
NC = Non-calculable  
RPD = Relative Percent Difference  
SPIKE = Fortified sample

**Validation Signature Page**

**Maxxam Job #: A878374**

---

The analytical data and all QC contained in this report were reviewed and validated by the following individual(s).

*Cristina Nervo*

---

CHRISTINA NERVO, Scientific Services

=====

Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.



6740 Campobello Rd  
Mississauga  
Ontario L5N 2L8

# REGULATION 243 DRINKING WATER SUBMISSION CHAIN OF CUSTODY RECORD

Tel: 905-817-5700  
Fax: 905-817-5779  
Toll Free: 1-800-563-6266

Samples are Regulated  
Under Reg 243

17-Jul-08 14:04  
JAMES ASPIN  
A878374  
CCO ENV-832  
Maxxam Analytics Inc.  
CAM FCD-00196/2

**Report to:**

Company Name: Dufferin-Peel Catholic District School Board  
Client Contact: Roland Lawton, General Manager of Physical Plant Facilities  
Address: 5685 Keaton Cres., Mississauga ON L5R 3H5  
E-Mail: [roland.lawton@dpcdsb.org](mailto:roland.lawton@dpcdsb.org) Tel: 905-890-0708 ext. 23204  
Project: Regulation 243 Fax: 905-890-5841  
Quote #: A83456 P.O.#: BN500636  
Lab Service Rep.: James Aspin

Invoice to:  Same as "Report to"

PAGE 01 OF 01

ANALYSIS REQUESTED

Specify School Type

M.O.E. Assigned Water System Number

School	Private School	Day nursery	
--------	----------------	-------------	--

Sample Identification	Sample Type (Please Specify Standing / Flushed)	Date Sampled	Time Sampled	Sampler Initials	Field Chlorine		No. of Bottles	Lead	School	Private School	Day nursery	M.O.E. Assigned Water System Number
					Total	Free						
Father Michael J. McGivney Kitchen - Staffroom	Standing	JUL 17	6:30A	JS			1	X	X			500012260
Kitchen - Staffroom	Flushed	JUL 17	7:07A	JS			1	X	X			

In accordance with the Safe Drinking Water Act (2002), notification of adverse results must be reported to the SAC, MOH & Waterworks. Failure to do so is an offence under the act. It is IMPERATIVE that the areas shaded in grey be filled out completely before submitting to Maxxam Analytics.

TAT (Turnaround Time)	Notification Information for School, Private Schools & Day Nurseries		LABORATORY USE ONLY
<b>RUSH TAT MUST HAVE PRIOR APPROVAL</b>	Facility Name: Father Michael J. McGivney	Medical Officer of Health	Received By: <i>MS FANGWANG</i>
<input checked="" type="checkbox"/> Normal (5-15 Working Days)	Facility Address: 450 Fernforest Drive	Region: Peel Region	Date: <i>08/07/17</i> Time: <i>14:04</i>
<input type="checkbox"/> RUSH (Specify Below)	Brampton, ON L6R 2P7	Tel #:	Comments: <i>10/10/10</i>
Rate Due:	Tel: 905-792-0630 Fax: 905-792-5054	Fax #:	
Lab #	E-mail:	Email:	<i>08 JUL 17 14:04</i>
Relinquished by (Sign Name): <i>Joe Sero</i>	Date: <i>17/07/08</i>	Laboratory Prepared Container:	<input type="checkbox"/> Yes <input type="checkbox"/> No
Relinquished by (Print Name): <i>JOE SERO</i>	Time: <i>7:15 AM</i>		

NOTE: Samples submitted to Maxxam Analytics must be received below 10°C. It is the Sampler's responsibility to ensure adequate cooling. Please ensure that a separate COC is completed.

SIF: Sample Inspection Resolved  By: \_\_\_\_\_ Page 6 of 6

DW METALS PRESERVATION