

Your P.O. #: BN500116
 Your Project #: IONA CATHOLIC S.S.
 Site:500013807
 Your C.O.C. #: na

Attention: Roland Lawton
 Dufferin Peel Catholic District...
 School Board
 40 Matheson Blvd W
 Mississauga, ON
 L5R 1C5

Report Date: 2008/09/22

CERTIFICATE OF ANALYSIS

MAXXAM JOB #: A8A4217
Received: 2008/09/12, 12:31

Sample Matrix: Water
 # Samples Received: 2

Analyses	Quantity	Date Extracted	Date Analyzed	Laboratory Method	Method Reference
Metals Analysis by ICPMS (as received) 0	2	2008/09/19	2008/09/19	CAM SOP-00447	EPA 6020

* RPDs calculated using raw data. The rounding of final results may result in the apparent difference.

(1) Metals analysis was performed on the sample 'as received'.

Encryption Key

Please direct all questions regarding this Certificate of Analysis to your Project Manager.

JAMES ASPIN, Project Manager
 Email: James.Aspin@maxxamAnalytics.com
 Phone# (905) 817-5700 Ext:5771

=====

Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.

For Service Group specific validation please refer to the Validation Signature Page

Total cover pages: 1

Maxxam Job #: A8A4217
 Report Date: 2008/09/22

Dufferin Peel Catholic District...
 Client Project #: IONA CATHOLIC S.S.
 Project name: 500013807
 Your P.O. #: BN500116

ELEMENTS BY ATOMIC SPECTROSCOPY (WATER)

Maxxam ID			AL9625	AL9626		
Sampling Date			2008/09/08 05:50	2008/09/08 06:28		
COC Number			na	na		
	Units	Criteria	IONA CATHOLIC S.S. KITCHEN- STAFFROOM STANDING	IONA CATHOLIC S.S. KITCHEN- STAFFROOM FLUSHED	RDL	QC Batch

Metals						
. Lead (Pb)	mg/L	0.01	0.0007	ND	0.0005	1619398

ND = Not detected
 RDL = Reportable Detection Limit
 QC Batch = Quality Control Batch
 CRITERIA: Ontario Regulation 243/07 made under the Safe Drinking Water Act, 2002.
 Released June 7, 2007

Maxxam Job #: A8A4217
Report Date: 2008/09/22

Dufferin Peel Catholic District...
Client Project #: IONA CATHOLIC S.S.
Project name: 500013807
Your P.O. #: BN500116

Package 1	14.0°C
-----------	--------

Each temperature is the average of up to three cooler temperatures taken at receipt

GENERAL COMMENTS

Results relate only to the items tested.

Dufferin Peel Catholic District...
 Attention: Roland Lawton
 Client Project #: IONA CATHOLIC S.S.
 P.O. #: BN500116
 Project name: 500013807

Quality Assurance Report
 Maxxam Job Number: MA8A4217

QA/QC Batch	QC Type	Parameter	Date Analyzed yyyy/mm/dd	Value	Recovery	Units	QC Limits
1619398 MIL	MATRIX SPIKE	. Lead (Pb)	2008/09/19		101	%	80 - 120
	Spiked Blank	. Lead (Pb)	2008/09/19		97	%	85 - 115
	Method Blank	. Lead (Pb)	2008/09/19	ND, RDL=0.0005		mg/L	
	RPD	. Lead (Pb)	2008/09/19	NC		%	25

ND = Not detected
 NC = Non-calculable
 RPD = Relative Percent Difference
 SPIKE = Fortified sample

Validation Signature Page

Maxxam Job #: A8A4217

The analytical data and all QC contained in this report were reviewed and validated by the following individual(s).



TROY CARRIERE, B.Sc., C.Chem, Scientific Specialist

=====

Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.



6740 Campobello Rd
Mississauga
Ontario L5N 2L8

12-Sep-08 12:31

JAMES ASPIN



A8A4217

NNA ENV-987

Toll Free: 1-800-563-6266

REGULATION 243 DRINKING WATER SUBMISSION CUSTODY RECORD

Samples are Regulated
Under Reg 243

Maxxam Analytics Inc.
CAM FCD-00196/2

Report to:

Company Name: Dufferin-Peel Catholic District School Board
 Client Contact: Roland Lawton, General Manager of Physical Plant Facilities
 Address: 5685 Keaton Cres., Mississauga ON L5R 3H5
 Mail: roland.lawton@dpcdsb.org Tel: 905-890-0708 ext. 23204
 Project: Regulation 243 Fax: 905-890-5841
 Quote #: A83456 P.O.#: BN500636
 Job Service Rep.: James Aspin

Invoice to: Same as "Report to"

PAGE ____ OF ____

ANALYSIS REQUESTED

Specify School Type
 M.O.E. Assigned Water System Number

Sample Identification	Sample Type (Please Specify Standing / Flushed)	Date Sampled	Time Sampled	Sampler Initials	Field Chlorine		No. of Bottles	Lead	School	Private School	Day nursery	M.O.E. Assigned Water System Number
					Total	Free						
Iona Catholic S.S. Kitchen - Staffroom	Standing	9-8-08	5:50				1	X	X			500013807
Iona Catholic S.S. Kitchen - Staffroom	Flushed	9-8-08	6:28				1	X	X			500013807

DW METALS PRESERVATION CHECKED

In accordance with the notification of adverse results must be reported to the SAC, MOH & Waterworks. Failure to do so is a violation of the Act. It is the responsibility of the sampler to ensure that the areas shaded in grey be filled out completely before submitting to Maxxam Analytics.

TAT (Turnaround Time)	Notification Information for School, Private Schools & Day Nurseries		LABORATORY USE ONLY
RUSH TAT MUST HAVE PRIOR APPROVAL	Facility Name: Iona Catholic S.S.	Medical Officer of Health	Received By: <i>[Signature]</i>
<input type="checkbox"/> Normal (5-15 Working Days)	Facility Address: 2170 South Sheridan Way	Region: Peel Region	Date: 08/09/12 Time: 12:31
<input type="checkbox"/> RUSH (Specify Below)	Mississauga, ON L5J 2M4		Comments: 14/10/11°C (ice pack safe)
Rate Due:	Tel: 905-823-0136 Fax: 905-823-0248	Tel #:	
Job #:	E-mail:	Fax #:	
Relinquished by (Sign Name): <i>[Signature]</i>	Date: <i>Sept. 8-08</i>	Email:	Laboratory Prepared Container: <input type="checkbox"/> Yes <input type="checkbox"/> No
Relinquished by (Print Name): SAMSON			

NOTE: Samples submitted at or below 10°C. If the sampling temperature is met. It is the Sampler's responsibility to ensure a separate container is used for waterworks. Please ensure that a separate container is used for waterworks.

SIF: Sample Inspection Resolved By: _____ Date: _____