

Your Project #: JOHN XXIII  
Site:500018669  
Your C.O.C. #: NA

**Attention: Roland Lawton**  
Dufferin Peel Catholic District...  
School Board  
40 Matheson Blvd W  
Mississauga, ON  
L5R 1C5

**Report Date: 2008/07/28**

**CERTIFICATE OF ANALYSIS**

**MAXXAM JOB #: A877970**  
**Received: 2008/07/17, 14:46**

Sample Matrix: Water  
# Samples Received: 2

Analyses	Quantity	Date Extracted	Date Analyzed	Laboratory Method	Method Reference
Metals Analysis by ICPMS (as received) 0	2	2008/07/25	2008/07/25	CAM SOP-00447	EPA 6020

\* RPDs calculated using raw data. The rounding of final results may result in the apparent difference.

(1) Metals analysis was performed on the sample 'as received'.

**Encryption Key**

Please direct all questions regarding this Certificate of Analysis to your Project Manager.

JAMES ASPIN, Project Manager  
Email: James.Aspin@maxxamAnalytics.com  
Phone# (905) 817-5700 Ext:5771

=====  
Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.

For Service Group specific validation please refer to the Validation Signature Page

Total cover pages: 1

Maxxam Job #: A877970  
Report Date: 2008/07/28

Dufferin Peel Catholic District...  
Client Project #: JOHN XXIII  
Project name: 500018669  
Sampler Initials: GP

**ELEMENTS BY ATOMIC SPECTROSCOPY (WATER)**

Maxxam ID			Z95425		
Sampling Date			2008/07/17 07:00		
COC Number			NA		
	<b>Units</b>	<b>Criteria</b>	<b>JOHN XX11</b>	<b>RDL</b>	<b>QC Batch</b>
			<b>KITCHEN-STAFFROOM/STANDING</b>		

<b>Metals</b>					
. Lead (Pb)	mg/L	<b>0.01</b>	0.0033	0.0005	1570400
<p>RDL = Reportable Detection Limit            QC Batch = Quality Control Batch            CRITERIA: Ontario Regulation 243/07 made under the Safe Drinking Water Act, 2002.            Released June 7, 2007</p>					

Maxxam ID			Z95426		
Sampling Date			2008/07/17 07:35		
COC Number			NA		
	<b>Units</b>	<b>Criteria</b>	<b>JOHN XX11</b>	<b>RDL</b>	<b>QC Batch</b>
			<b>KITCHEN-STAFFROOM/FLUSHED</b>		

<b>Metals</b>					
. Lead (Pb)	mg/L	<b>0.01</b>	0.0006	0.0005	1570417
<p>RDL = Reportable Detection Limit            QC Batch = Quality Control Batch            CRITERIA: Ontario Regulation 243/07 made under the Safe Drinking Water Act, 2002.            Released June 7, 2007</p>					

Maxxam Job #: A877970  
Report Date: 2008/07/28

Dufferin Peel Catholic District...  
Client Project #: JOHN XXIII  
Project name: 500018669  
Sampler Initials: GP

Package 1	12.3°C
-----------	--------

Each temperature is the average of up to three cooler temperatures taken at receipt

**GENERAL COMMENTS**

**Results relate only to the items tested.**

Dufferin Peel Catholic District...  
Attention: Roland Lawton  
Client Project #: JOHN XXIII  
P.O. #:  
Project name: 500018669

Quality Assurance Report  
Maxxam Job Number: MA877970

QA/QC Batch	QC Type	Parameter	Date Analyzed yyyy/mm/dd	Value	Recovery	Units	QC Limits
1570400 HRE	MATRIX SPIKE	. Lead (Pb)	2008/07/25		100	%	80 - 120
	Spiked Blank	. Lead (Pb)	2008/07/25		101	%	85 - 115
	Method Blank	. Lead (Pb)	2008/07/25	ND, RDL=0.5		ug/L	
	RPD	. Lead (Pb)	2008/07/25	NC		%	25
1570417 HRE	MATRIX SPIKE	. Lead (Pb)	2008/07/25		100	%	80 - 120
	Spiked Blank	. Lead (Pb)	2008/07/25		97	%	85 - 115
	Method Blank	. Lead (Pb)	2008/07/25	ND, RDL=0.5		ug/L	
	RPD	. Lead (Pb)	2008/07/25	1.2		%	25

ND = Not detected  
 NC = Non-calculable  
 RPD = Relative Percent Difference  
 SPIKE = Fortified sample

**Validation Signature Page**

**Maxxam Job #: A877970**

---

The analytical data and all QC contained in this report were reviewed and validated by the following individual(s).

*Cristina Nervo*

---

CHRISTINA NERVO, Scientific Services

=====

Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.



6740 Campobello Rd  
Mississauga  
Ontario L5N 2L8

REGULATION 243 DRINKING WATER SU  
**CHAIN OF CUSTODY RECORD**

Tel: 905-817-5700  
Fax: 905-817-5779  
Toll Free: 1-800-563-6266

Samples are Regulated  
Under Reg 243

17-Jul-08 14:46  
JAMES ASPIN  
A877970  
FW ENV-756  
CAM FCD-00196/2

**Report to:**

Company Name: Dufferin-Peel Catholic District School Board  
Client Contact: Roland Lawton, General Manager of Physical Plant Facilities  
Address: 5685 Keaton Cres., Mississauga ON L5R 3H5  
E-Mail: roland.lawton@dpcdsb.org Tel: 905-890-0708 ext. 23204  
Project: Regulation 243 Fax: 905-890-5841  
Quote #: A83456 P.O.#: BN500636  
Lab Service Rep.: James Aspen

Invoice to:  Same as "Report to"

DW METALS  
PRESERVATION  
CHECKED

PAGE \_\_\_\_ OF \_\_\_\_

ANALYSIS REQUESTED

Specify School Type

M.O.E. Assigned Water System Number

School	Private School	Day nursery	
			500018669

Sample Identification	Sample Type (Please Specify Standing / Flushed)	Date Sampled	Time Sampled	Sampler Initials	Field Chlorine		No. of Bottles	Lead	School	Private School	Day nursery
					Total	Free					
John XX111											
Kitchen - Staffroom	Standing	7-17-08	7:00 AM				1	X	X		
Kitchen - Staffroom	Flushed	7-17-08	7:35 AM				1	X	X		

SIF : Sample Inspection  
Resolved  By: \_\_\_\_\_  
Date: \_\_\_\_\_

\* In accordance with the Safe Drinking Water Act (2002), notification of adverse results must be reported to the Ministry of Health. Failure to do so is an offence under the act. It is IMPERATIVE that the areas shaded in grey be filled in.

TAT (Turnaround Time)	Notification Information for School, Private Schools & Day Nurseries	LABORATORY USE ONLY
<b>RUSH TAT MUST HAVE PRIOR APPROVAL</b>	Facility Name: John XX111 Facility Address: 915 McBride Avenue Mississauga, ON L5C 1M1	Received By: <i>J. James Wong</i> Date: 08/07/17 Time: 14:46 Comments: B/D/1/13
<input checked="" type="checkbox"/> Normal (5-15 Working Days)	Tel #: _____ Fax #: _____	
<input type="checkbox"/> RUSH (Specify Below)	E-mail: _____	
Date Due:		JUL 17 '08 14:46
PN #		
Relinquished by (Sign Name): <i>Generoso Pineda</i>	Date: JULY 17, 2008	Laboratory Prepared Container:
Relinquished by (Print Name): GENEROSO PINEDA	Time: 7:00 A.M. 7:35 A.M.	<input type="checkbox"/> Yes <input type="checkbox"/> No

NOTE: Samples submitted to Maxxam Analytics must be received below 10°C.

It is the Sampler's responsibility to ensure adequate cooling is provided so that the receiving temperature is met.  
Please ensure that a separate COC document is used for each individual waterworks.