

Your Project #: MARY FIX CATHOLIC  
Site:500018835  
Your C.O.C. #: NA

**Attention: Roland Lawton**  
Dufferin Peel Catholic District...  
School Board  
40 Matheson Blvd W  
Mississauga, ON  
L5R 1C5

**Report Date: 2008/07/28**

**CERTIFICATE OF ANALYSIS**

**MAXXAM JOB #: A877969**  
**Received: 2008/07/17, 14:56**

Sample Matrix: Water  
# Samples Received: 2

Analyses	Quantity	Date Extracted	Date Analyzed	Laboratory Method	Method Reference
Metals Analysis by ICPMS (as received) 0	2	2008/07/25	2008/07/25	CAM SOP-00447	EPA 6020

\* RPDs calculated using raw data. The rounding of final results may result in the apparent difference.

(1) Metals analysis was performed on the sample 'as received'.

**Encryption Key**

Please direct all questions regarding this Certificate of Analysis to your Project Manager.

JAMES ASPIN, Project Manager  
Email: James.Aspin@maxxamAnalytics.com  
Phone# (905) 817-5700 Ext:5771

=====  
Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.

For Service Group specific validation please refer to the Validation Signature Page

Total cover pages: 1

Maxxam Job #: A877969  
Report Date: 2008/07/28

Dufferin Peel Catholic District...  
Client Project #: MARY FIX CATHOLIC  
Project name: 500018835  
Sampler Initials: DV

**ELEMENTS BY ATOMIC SPECTROSCOPY (WATER)**

Maxxam ID			Z95423		
Sampling Date			2008/07/17 07:10		
COC Number			NA		
	<b>Units</b>	<b>Criteria</b>	<b>MARY FIX CATHOLIC KITCHEN-STAFFROOM/STANDING</b>	<b>RDL</b>	<b>QC Batch</b>

<b>Metals</b>					
. Lead (Pb)	mg/L	<b>0.01</b>	0.0007	0.0005	1570417
<p>RDL = Reportable Detection Limit            QC Batch = Quality Control Batch            CRITERIA: Ontario Regulation 243/07 made under the Safe Drinking Water Act, 2002.            Released June 7, 2007</p>					

Maxxam ID			Z95424		
Sampling Date			2008/07/17 07:45		
COC Number			NA		
	<b>Units</b>	<b>Criteria</b>	<b>MARY FIX CATHOLIC KITCHEN-STAFFROOM/FLUSHED</b>	<b>RDL</b>	<b>QC Batch</b>

<b>Metals</b>					
. Lead (Pb)	mg/L	<b>0.01</b>	ND	0.0005	1570400
<p>ND = Not detected            RDL = Reportable Detection Limit            QC Batch = Quality Control Batch            CRITERIA: Ontario Regulation 243/07 made under the Safe Drinking Water Act, 2002.            Released June 7, 2007</p>					

Maxxam Job #: A877969  
Report Date: 2008/07/28

Dufferin Peel Catholic District...  
Client Project #: MARY FIX CATHOLIC  
Project name: 500018835  
Sampler Initials: DV

Package 1	13.3°C
-----------	--------

Each temperature is the average of up to three cooler temperatures taken at receipt

**GENERAL COMMENTS**

**Results relate only to the items tested.**

Dufferin Peel Catholic District...  
Attention: Roland Lawton  
Client Project #: MARY FIX CATHOLIC  
P.O. #:  
Project name: 500018835

Quality Assurance Report  
Maxxam Job Number: MA877969

QA/QC Batch	QC Type	Parameter	Date Analyzed yyyy/mm/dd	Value	Recovery	Units	QC Limits
1570400 HRE	MATRIX SPIKE	. Lead (Pb)	2008/07/25		100	%	80 - 120
	Spiked Blank	. Lead (Pb)	2008/07/25		101	%	85 - 115
	Method Blank	. Lead (Pb)	2008/07/25	ND, RDL=0.5		ug/L	
	RPD	. Lead (Pb)	2008/07/25	NC		%	25
1570417 HRE	MATRIX SPIKE	. Lead (Pb)	2008/07/25		100	%	80 - 120
	Spiked Blank	. Lead (Pb)	2008/07/25		97	%	85 - 115
	Method Blank	. Lead (Pb)	2008/07/25	ND, RDL=0.5		ug/L	
	RPD	. Lead (Pb)	2008/07/25	1.2		%	25

ND = Not detected  
NC = Non-calculable  
RPD = Relative Percent Difference  
SPIKE = Fortified sample

**Validation Signature Page**

**Maxxam Job #: A877969**

---

The analytical data and all QC contained in this report were reviewed and validated by the following individual(s).

*Cristina Nervo*

---

CHRISTINA NERVO, Scientific Services

=====

Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.



6740 Campobello Rd  
Mississauga  
Ontario L5N 2L8

# REGULATION 243 DRINKING WATER CHAIN OF CUSTODY RECORD

Tel: 905-817-5700  
Fax: 905-817-5779  
Toll Free: 1-800-563-6266

Samples are Regulated  
Under Reg 243

17-Jul-08 14:56  
JAMES ASPIN  
A877969  
FW ENV-756  
CAM FCD-00196/2

**Report to:**

Company Name: Dufferin-Peel Catholic District School Board  
Client Contact: Roland Lawton, General Manager of Physical Plant Facilities  
Address: 5685 Keaton Cres., Mississauga ON L5R 3H5  
E-Mail: [roland.lawton@dpcdsb.org](mailto:roland.lawton@dpcdsb.org) Tel: 905-890-0708 ext. 23204  
Project: Regulation 243 Fax: 905-890-5841  
Quote #: A83456 P.O.#: BN500636  
Lab Service Rep.: James Aspin

Invoice to:  Same as "Report to"

DW METALS  
PRESERVATION  
CHECKED

PAGE \_\_\_\_ OF \_\_\_\_

**ANALYSIS REQUESTED**

Specify School Type

M.O.E. Assigned Water System Number

School	Private School	Day nursery
<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>

500018825

Sample Identification	Sample Type (Please Specify Standing / Flushed)	Date Sampled	Time Sampled	Sampler Initials	Field Chlorine		No. of Bottles	Lead	School	Private School	Day nursery
					Total	Free					
Mary Fix Catholic Kitchen - Staffroom	Standing	17-7-08	7:10 AM	SV			1	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Mary Fix Catholic Kitchen - Staffroom	Flushed	17-7-08	7:45 AM	SV			1	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	

SIF : Sample Inspection

Resolved  By: \_\_\_\_\_

Date: \_\_\_\_\_ in Analytics.

In accordance with the Safe Drinking Water Act (2002), notification of adverse results must be provided to the school. Failure to do so is an offence under the act. It is **IMPERATIVE** that the areas shaded in grey below be completed.

TAT (Turnaround Time)	Notification Information for School, Private Schools & Day Nurseries	LABORATORY USE ONLY
<b>RUSH TAT MUST HAVE PRIOR APPROVAL</b>	Facility Name: Mary Fix Catholic Facility Address: 486 Paisley Blvd., West Mississauga, ON L5B 2M4	Received By: <i>ZORNA ZENETA</i> Date: <i>08/07/08</i> Time: <i>14:56</i> Comments:
<input type="checkbox"/> Normal (5-15 Working Days)		
<input type="checkbox"/> RUSH (Specify Below)		
Rate Due:	Tel: 905-270-3140 Fax: 905-270-3475	
Lab #	E-mail:	
Relinquished by (Sign Name): <i>SV</i>	Date: <i>17-7-08</i>	Laboratory Prepared Container:
Relinquished by (Print Name): <i>D1210 VITO</i>	Time: <i>7:10 AM</i>	<input type="checkbox"/> Yes <input type="checkbox"/> No

**NOTE:** Samples submitted to Maxxam Analytics must be received below 10°C.

It is the Sampler's responsibility to ensure adequate cooling is provided so that this Page 1 of 1 temperature is met. Please ensure that a separate COC document is used for each individual water sample.

12/14/14  
08 JUL 17 14:56