

Your P.O. #: BN500116
Your Project #: OUT LADY OF MOUNT CARMEL
Site:500018916
Your C.O.C. #: NA

Attention: Roland Lawton
Dufferin Peel Catholic District...
School Board
40 Matheson Blvd W
Mississauga, ON
L5R 1C5

Report Date: 2008/08/08

CERTIFICATE OF ANALYSIS

MAXXAM JOB #: A885283
Received: 2008/08/01, 13:05

Sample Matrix: Water
Samples Received: 2

Analyses	Quantity	Date Extracted	Date Analyzed	Laboratory Method	Method Reference
Metals Analysis by ICPMS (as received) ①	1	2008/08/06	2008/08/07	CAM SOP-00447	EPA 6020
Metals Analysis by ICPMS (as received) ①	1	2008/08/07	2008/08/07	CAM SOP-00447	EPA 6020

* RPDs calculated using raw data. The rounding of final results may result in the apparent difference.

(1) Metals analysis was performed on the sample 'as received'.

Encryption Key

Please direct all questions regarding this Certificate of Analysis to your Project Manager.

JAMES ASPIN, Project Manager
Email: James.Aspin@maxxamAnalytics.com
Phone# (905) 817-5700 Ext:5771

=====
Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.

For Service Group specific validation please refer to the Validation Signature Page

Total cover pages: 1

Maxxam Job #: A885283
Report Date: 2008/08/08

Dufferin Peel Catholic District...
Client Project #: OUT LADY OF MOUNT CARMEL
Project name: 500018916
Your P.O. #: BN500116

ELEMENTS BY ATOMIC SPECTROSCOPY (WATER)

Maxxam ID			AC8058		AC8059		
Sampling Date			2008/08/01 06:40		2008/08/01 07:15		
COC Number			NA		NA		
	Units	Criteria	OUR LADY OF MOUNT CARMEL STANDING	QC Batch	OUR LADY OF MOUNT CARMEL FLUSHED	RDL	QC Batch

Metals							
. Lead (Pb)	mg/L	0.01	0.0014	1579822	ND	0.0005	1580482

ND = Not detected
RDL = Reportable Detection Limit
QC Batch = Quality Control Batch
CRITERIA: Ontario Regulation 243/07 made under the Safe Drinking Water Act, 2002.

Released June 7, 2007

Maxxam Job #: A885283
Report Date: 2008/08/08

Dufferin Peel Catholic District...
Client Project #: OUT LADY OF MOUNT CARMEL
Project name: 500018916
Your P.O. #: BN500116

Package 1	11.0°C
-----------	--------

Each temperature is the average of up to three cooler temperatures taken at receipt

GENERAL COMMENTS

Results relate only to the items tested.

Dufferin Peel Catholic District...
Attention: Roland Lawton
Client Project #: OUT LADY OF MOUNT CARMEL
P.O. #: BN500116
Project name: 500018916

Quality Assurance Report
Maxxam Job Number: MA885283

QA/QC Batch	QC Type	Parameter	Date Analyzed yyyy/mm/dd	Value	Recovery	Units	QC Limits
1579822 MIL	MATRIX SPIKE	. Lead (Pb)	2008/08/07		102	%	80 - 120
	Spiked Blank	. Lead (Pb)	2008/08/07		100	%	85 - 115
	Method Blank	. Lead (Pb)	2008/08/07	ND, RDL=0.0005		mg/L	
	RPD	. Lead (Pb)	2008/08/07	NC		%	25
1580482 HRE	MATRIX SPIKE	. Lead (Pb)	2008/08/07		99	%	80 - 120
	Spiked Blank	. Lead (Pb)	2008/08/07		101	%	85 - 115
	Method Blank	. Lead (Pb)	2008/08/07	ND, RDL=0.0005		mg/L	
	RPD	. Lead (Pb)	2008/08/07	NC		%	25

ND = Not detected
 NC = Non-calculable
 RPD = Relative Percent Difference
 SPIKE = Fortified sample

Validation Signature Page

Maxxam Job #: A885283

The analytical data and all QC contained in this report were reviewed and validated by the following individual(s).



TROY CARRIERE, B.Sc., C.Chem, Scientific Specialist

Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.



6740 Campobello Rd
Mississauga
Ontario L5N 2L8

REGULATION 243 DRINKING WATER SUBM CHAIN OF CUSTODY RECORD

Tel: 905-817-5700
Fax: 905-817-5779
Toll Free: 1-800-563-6266

Samples are Regulated
Under Reg 243

1-Aug-08 13:05
JAMES ASPIN
A885283
M_O ENV-775

Report to:

Company Name: Dufferin-Peel Catholic District School Board
Client Contact: Roland Lawton, General Manager of Physical Plant Facilities
Address: 5685 Keaton Cres., Mississauga ON L5R 3H5
E-Mail: roland.lawton@dpcdsb.org Tel: 905-890-0708 ext. 23204
Project: Regulation 243 Fax: 905-890-5841
Quote #: A83456 P.O.#: BN500636
Lab Service Rep.: James Aspen

Invoice to: Same as "Report to"

PAGE ____ OF ____

ANALYSIS REQUESTED

Specify School Type

M.O.E. Assigned Water System Number

500018916

Lead	School	Private School	Day nursery
	<input checked="" type="checkbox"/>		
	<input checked="" type="checkbox"/>		

Sample Identification	Sample Type (Please Specify Standing / Flushed)	Date Sampled	Time Sampled	Sampler Initials	Field Chlorine		No. of Bottles	Lead
					Total	Free		
Our Lady of Mount Carmel Kitchen - Staffroom	Standing	AUG 1	6:40 AM	C.K			1	X
Kitchen - Staffroom	Flushed	AUG 1	7:15 AM	C.K			1	X

* In accordance with the Safe Drinking Water Act (2002), notification of adverse results must be reported to the SAC, MOH & Waterworks. Failure to do so is an offence under the act. It is **IMPERATIVE** that the areas shaded in grey be filled out completely before submitting to Maxxam Analytics.

TAT (Turnaround Time)	Notification Information for School, Private Schools & Day Nurseries		LABORATORY USE ONLY	
RUSH TAT MUST HAVE PRIOR APPROVAL	Facility Name: Our Lady of Mount Carmel	Medical Officer of Health	Received By: <i>James Aspin</i>	
<input checked="" type="checkbox"/> Normal (5-15 Working Days)	Facility Address: 3700 Trelawny Circle	Region: Peel Region	Date: 08/08/08	Time: 13:05
<input type="checkbox"/> RUSH (Specify Below)	Mississauga, ON L5N 5J7	Tel #:	Comments: 08 AUG 1 13:05	
Date Due:	Tel: 905-824-1025 Fax: 905-824-4549	Fax #:	12/11/10	
PN #	E-mail:	Email:		
Relinquished by (Sign Name):	Date:	Laboratory Prepared Container: <input type="checkbox"/> Yes <input type="checkbox"/> No		
Relinquished by (Print Name):				

NO : Sample Inspection
It is the Sampler's responsibility to ensure that samples are properly collected and preserved. Please
Resolved By: _____
Date: _____

DW METALS
PRESERVATION
CHECKED