

Your P.O. #: BN500636
Your Project #: ST.ALBERT OF JERUSALEM
Site:500019033
Your C.O.C. #: na

Attention: Roland Lawton
Dufferin Peel Catholic District...
School Board
40 Matheson Blvd W
Mississauga, ON
L5R 1C5

Report Date: 2008/09/22

CERTIFICATE OF ANALYSIS

MAXXAM JOB #: A8A4214
Received: 2008/09/12, 12:32

Sample Matrix: Water
Samples Received: 2

Analyses	Quantity	Date Extracted	Date Analyzed	Laboratory Method	Method Reference
Metals Analysis by ICPMS (as received) 0	2	2008/09/19	2008/09/19	CAM SOP-00447	EPA 6020

* RPDs calculated using raw data. The rounding of final results may result in the apparent difference.

(1) Metals analysis was performed on the sample 'as received'.

Encryption Key

Please direct all questions regarding this Certificate of Analysis to your Project Manager.

JAMES ASPIN, Project Manager
Email: James.Aspin@maxxamAnalytics.com
Phone# (905) 817-5700 Ext:5771

=====
Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.

For Service Group specific validation please refer to the Validation Signature Page

Total cover pages: 1

Maxxam Job #: A8A4214
 Report Date: 2008/09/22

Dufferin Peel Catholic District...
 Client Project #: ST.ALBERT OF JERUSALEM
 Project name: 500019033
 Your P.O. #: BN500636
 Sampler Initials: PS

ELEMENTS BY ATOMIC SPECTROSCOPY (WATER)

Maxxam ID			AL9609	AL9610		
Sampling Date			2008/09/12 06:30	2008/09/12 07:05		
COC Number			na	na		
	Units	Criteria	ST ALBERT OF JERUSALEM STANDING	ST ALBERT OF JERUSALEM FLUSHED	RDL	QC Batch

Metals						
Lead (Pb)	mg/L	0.01	0.0014	0.0012	0.0005	1619348

RDL = Reportable Detection Limit
 QC Batch = Quality Control Batch
 CRITERIA: Ontario Regulation 243/07 made under the Safe Drinking Water Act, 2002.
 Released June 7, 2007

Maxxam Job #: A8A4214
Report Date: 2008/09/22

Dufferin Peel Catholic District...
Client Project #: ST.ALBERT OF JERUSALEM
Project name: 500019033
Your P.O. #: BN500636
Sampler Initials: PS

Package 1	12.0°C
-----------	--------

Each temperature is the average of up to three cooler temperatures taken at receipt

GENERAL COMMENTS

Results relate only to the items tested.

Dufferin Peel Catholic District...
 Attention: Roland Lawton
 Client Project #: ST.ALBERT OF JERUSALEM
 P.O. #: BN500636
 Project name: 500019033

Quality Assurance Report
 Maxxam Job Number: MA8A4214



QA/QC Batch	QC Type	Parameter	Date Analyzed yyyy/mm/dd	Value	Recovery	Units	QC Limits
1619348 JBW	MATRIX SPIKE	. Lead (Pb)	2008/09/19		105	%	80 - 120
	Spiked Blank	. Lead (Pb)	2008/09/19		100	%	85 - 115
	Method Blank	. Lead (Pb)	2008/09/19	ND, RDL=0.0005		mg/L	
	RPD	. Lead (Pb)	2008/09/19	NC		%	25

ND = Not detected
 NC = Non-calculable
 RPD = Relative Percent Difference
 SPIKE = Fortified sample

Validation Signature Page

Maxxam Job #: A8A4214

The analytical data and all QC contained in this report were reviewed and validated by the following individual(s).



TROY CARRIERE, B.Sc., C.Chem, Scientific Specialist

=====

Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.



OF CUSTODY RECORD

Samples are Regulated
Under Reg 243

Report to:

Company Name: Dufferin-Peel Catholic District School Board
Client Contact: Roland Lawton, General Manager of Physical Plant Facilities
Address: 5685 Keaton Cres., Mississauga ON L5R 3H5
E-Mail: roland.lawton@dpcdsb.org Tel: 905-890-0708 ext. 23204
Project: Regulation 243 Fax: 905-890-5841
Quote #: A83456 P.O.#: BN500636
Lab Service Rep.: James Aspin

Invoice to: Same as "Report to"

PAGE ____ OF ____

ANALYSIS REQUESTED

Specify School Type	M.O.E. Assigned Water System Number
School	
Private School / Day nursery	

500019033

08 SEP 12 12:32

Sample Identification	Sample Type (Please Specify Standing / Flushed)	Date Sampled	Time Sampled	Sampler Initials	Field Chlorine		No. of Bottles	Lead	School	Private School	Day nursery
					Total	Free					
St. Albert of Jerusalem											
Kitchen - Staffroom	Standing		6:30 AM				1	X	X		
Kitchen - Staffroom	Flushed		7:05 AM				1	X	X		

DW METALS
PRESERVATION
CHECKED

2002), notification of adverse results must be reported to the SAC, MOH & Waterworks.
IMPERATIVE that the areas shaded in grey be filled out completely before submitting to Maxxam Analytics.

TAT (Turnaround Time)	Notification Information for School, Private Schools & Day Nurseries		LABORATORY USE ONLY
RUSH TAT MUST HAVE PRIOR APPROVAL	Facility Name: St. Albert of Jerusalem	Medical Officer of Health	Received By: <i>Nagla Mikhail</i>
<input checked="" type="checkbox"/> Normal (5-15 Working Days)	Facility Address: 7185 Rosehurst Drive	Region: Peel Region	Date: <i>08/09/12</i> Time: <i>12:32</i>
<input type="checkbox"/> RUSH (Specify Below)	Mississauga, ON L5N 7G6	Tel #:	Comments: <i>12/12/12⁰ insufficient cooling</i>
Date Due:	Tel: 905-785-9298 Fax: 905-785-1227	Fax #:	
PN #	E-mail:	Email:	
Relinquished by (Sign Name): <i>Paulo Manuel Silva</i>	Date: <i>12/19/08</i>	Laboratory Prepared Container:	
Relinquished by (Print Name): <i>PAULO MANUEL SILVA</i>	Time: <i>7:05 AM</i>	<input type="checkbox"/> Yes <input type="checkbox"/> No	

NOTE: Samples submitted to Maxxam Analytics must be received by 12:00 PM.
It is the Sampler's responsibility to ensure adequate cooling is provided so that the sample is preserved.
Please ensure that a separate COC document is used for SIF: Sample Inspection

Resolved By: _____
Date: _____