

Your P.O. #: BN500116
Your Project #: ST EDITH STEIN
Site:500019085
Your C.O.C. #: na

Attention: Roland Lawton
Dufferin Peel Catholic District...
School Board
40 Matheson Blvd W
Mississauga, ON
L5R 1C5

Report Date: 2008/07/25

CERTIFICATE OF ANALYSIS

MAXXAM JOB #: A878387
Received: 2008/07/17, 14:08

Sample Matrix: Water
Samples Received: 2

Analyses	Quantity	Date Extracted	Date Analyzed	Laboratory Method	Method Reference
Metals Analysis by ICPMS (as received) ①	1	2008/07/24	2008/07/24	CAM SOP-00447	EPA 6020
Metals Analysis by ICPMS (as received) ①	1	2008/07/24	2008/07/25	CAM SOP-00447	EPA 6020

* RPDs calculated using raw data. The rounding of final results may result in the apparent difference.

(1) Metals analysis was performed on the sample 'as received'.

Encryption Key

Please direct all questions regarding this Certificate of Analysis to your Project Manager.

JAMES ASPIN, Project Manager
Email: James.Aspin@maxxamAnalytics.com
Phone# (905) 817-5700 Ext:5771

=====
Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.

For Service Group specific validation please refer to the Validation Signature Page

Total cover pages: 1

Maxxam Job #: A878387
Report Date: 2008/07/25

Dufferin Peel Catholic District...
Client Project #: ST EDITH STEIN
Project name: 500019085
Your P.O. #: BN500116
Sampler Initials: LC

ELEMENTS BY ATOMIC SPECTROSCOPY (WATER)

Maxxam ID			Z96920		Z96921		
Sampling Date			2008/07/15 07:20		2008/07/15 07:55		
COC Number			na		na		
	Units	Criteria	ST. EDITH STEIN KITCHEN- STAFFROOM JULY 15 7:20 STANDING	QC Batch	ST. EDITH STEIN KITCHEN- STAFFROOM JULY 15 7:55 FLUSHED	RDL	QC Batch

Metals							
. Lead (Pb)	mg/L	0.01	0.0021	1569253	0.0026	0.0005	1569634

RDL = Reportable Detection Limit
QC Batch = Quality Control Batch
CRITERIA: Ontario Regulation 243/07 made under the Safe Drinking Water Act, 2002.

Released June 7, 2007

Maxxam Job #: A878387
Report Date: 2008/07/25

Dufferin Peel Catholic District...
Client Project #: ST EDITH STEIN
Project name: 500019085
Your P.O. #: BN500116
Sampler Initials: LC

Package 1	8.3°C
-----------	-------

Each temperature is the average of up to three cooler temperatures taken at receipt

GENERAL COMMENTS

Results relate only to the items tested.

Dufferin Peel Catholic District...
Attention: Roland Lawton
Client Project #: ST EDITH STEIN
P.O. #: BN500116
Project name: 500019085

Quality Assurance Report
Maxxam Job Number: MA878387

QA/QC Batch	QC Type	Parameter	Date Analyzed yyyy/mm/dd	Value	Recovery	Units	QC Limits
1569253 JBW	MATRIX SPIKE	. Lead (Pb)	2008/07/24		106	%	80 - 120
	Spiked Blank	. Lead (Pb)	2008/07/24		99	%	85 - 115
	Method Blank	. Lead (Pb)	2008/07/24	ND, RDL=0.5		ug/L	
	RPD	. Lead (Pb)	2008/07/24	NC		%	25
1569634 HRE	MATRIX SPIKE	. Lead (Pb)	2008/07/25		98	%	80 - 120
	Spiked Blank	. Lead (Pb)	2008/07/25		102	%	85 - 115
	Method Blank	. Lead (Pb)	2008/07/25	ND, RDL=0.5		ug/L	
	RPD	. Lead (Pb)	2008/07/25	NC		%	25

ND = Not detected
 NC = Non-calculable
 RPD = Relative Percent Difference
 SPIKE = Fortified sample

Validation Signature Page

Maxxam Job #: A878387

The analytical data and all QC contained in this report were reviewed and validated by the following individual(s).

Cristina Nervo

CHRISTINA NERVO, Scientific Services

=====

Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.

SIF : Sample Inspection

M

Resolved By: _____

Date: _____

6740 Campobello Rd
Mississauga
Ontario L5N 2L8

**REGI
CH**

Tel: 905-817-5777
Fax: 905-817-5779
Toll Free: 1-800-563-6266

17-Jul-08 14:08

JAMES ASPIN



A878387

NNA

ENV-998

**WATER S
RECO**

**DW METALS
PRESERVATION
CHECKED**

**Campuses are Regulated
Under Reg 243**

Maxxam Analytics Inc.
CAM FCD-00196/2

Report to:

Company Name:	Dufferin-Peel Catholic District School Board		
Client Contact:	Roland Lawton, General Manager of Physical Plant Facilities		
Address:	5685 Keaton Cres., Mississauga ON L5R 3H5		
E-Mail:	<u>roland.lawton@dpcdsb.org</u>	Tel:	905-890-0708 ext. 23204
Project:	Regulation 243	Fax:	905-890-5841
Quote #:	A83456	P.O.#:	BN500636
Lab Service Rep.:	James Aspen		

Invoice to: Same as "Report to"

PAGE ____ OF ____

ANALYSIS REQUESTED

Specify
School
Type

M.O.E. Assigned
Water System
Number

Sample Identification	Sample Type (Please Specify Standing / Flushed)	Date Sampled	Time Sampled	Sampler Initials	Field Chlorine		No. of Bottles	Lead	School	Private School	Day nursery	M.O.E. Assigned Water System Number
					Total	Free						
St. Edith Stein												
Kitchen - Staffroom	Standing	July 15, 08	7:20 AM	M			1	x	x			500019085
Kitchen - Staffroom	Flushed	July 15, 08	7:55 AM	M			1	x	x			

* In accordance with the Safe Drinking Water Act (2002), notification of adverse results must be reported to the SAC, MOH & Waterworks. Failure to do so is an offence under the act. **It is IMPERATIVE** that the areas shaded in grey be filled out completely before submitting to Maxxam Analytics.

TAT (Turnaround Time)	Notification Information for School, Private Schools & Day Nurseries		LABORATORY USE ONLY
RUSH TAT MUST HAVE PRIOR APPROVAL	Facility Name: St. Edith Stein	Medical Officer of Health	Received By: <i>[Signature]</i>
<input checked="" type="checkbox"/> Normal (5-15 Working Days)	Facility Address: 6234 Osprey Blvd., Mississauga, ON L5N 5V5	Region: Peel Region	Date: 08/07/08 Time: 14:08
<input type="checkbox"/> RUSH (Specify Below)		Tel #:	Comments: 81819
Date Due:	Tel: 905-824-5777 Fax: 905-824-3871	Fax #:	08 JUL 17 14:08
PN #	E-mail:	Email:	
Relinquished by (Sign Name): <i>[Signature]</i>	Date: July 17/08	Laboratory Prepared Container:	<input type="checkbox"/> Yes <input type="checkbox"/> No
Relinquished by (Print Name): Lydia Chatoff	Time: July 2:45 pm		

NOTE: Samples submitted to Maxxam Analytics must be received below 10°C.

It is the Sampler's responsibility to ensure adequate cooling is provided so that this receiving temperature is met.

Please ensure that a separate COC document is used for each indiv