

Your Project #: ST GERARD
Site:500018734
Your C.O.C. #: NA

Attention: Roland Lawton
Dufferin Peel Catholic District...
School Board
40 Matheson Blvd W
Mississauga, ON
L5R 1C5

Report Date: 2008/07/28

CERTIFICATE OF ANALYSIS

MAXXAM JOB #: A877972
Received: 2008/07/17, 14:46

Sample Matrix: Water
Samples Received: 2

Analyses	Quantity	Date Extracted	Date Analyzed	Laboratory Method	Method Reference
Metals Analysis by ICPMS (as received) 0	2	2008/07/25	2008/07/25	CAM SOP-00447	EPA 6020

* RPDs calculated using raw data. The rounding of final results may result in the apparent difference.

(1) Metals analysis was performed on the sample 'as received'.

Encryption Key

Please direct all questions regarding this Certificate of Analysis to your Project Manager.

JAMES ASPIN, Project Manager
Email: James.Aspin@maxxamAnalytics.com
Phone# (905) 817-5700 Ext:5771

=====
Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.

For Service Group specific validation please refer to the Validation Signature Page

Total cover pages: 1

Maxxam Job #: A877972
Report Date: 2008/07/28

Dufferin Peel Catholic District...
Client Project #: ST GERARD
Project name: 500018734
Sampler Initials: JJ

ELEMENTS BY ATOMIC SPECTROSCOPY (WATER)

Maxxam ID			Z95430		
Sampling Date			2008/07/15 06:25		
COC Number			NA		
	Units	Criteria	ST. GERARD KITCHEN-STAFFROOM/STANDING	RDL	QC Batch

Metals					
. Lead (Pb)	mg/L	0.01	0.0010	0.0005	1570400
<p>RDL = Reportable Detection Limit QC Batch = Quality Control Batch CRITERIA: Ontario Regulation 243/07 made under the Safe Drinking Water Act, 2002. Released June 7, 2007</p>					

Maxxam ID			Z95431		
Sampling Date			2008/07/15 07:05		
COC Number			NA		
	Units	Criteria	ST. GERARD KITCHEN-STAFFROOM/FLUSHED	RDL	QC Batch

Metals					
. Lead (Pb)	mg/L	0.01	ND	0.0005	1570417
<p>ND = Not detected RDL = Reportable Detection Limit QC Batch = Quality Control Batch CRITERIA: Ontario Regulation 243/07 made under the Safe Drinking Water Act, 2002. Released June 7, 2007</p>					

Maxxam Job #: A877972
Report Date: 2008/07/28

Dufferin Peel Catholic District...
Client Project #: ST GERARD
Project name: 500018734
Sampler Initials: JJ

Package 1	13.0°C
-----------	--------

Each temperature is the average of up to three cooler temperatures taken at receipt

GENERAL COMMENTS

Results relate only to the items tested.

Dufferin Peel Catholic District...
Attention: Roland Lawton
Client Project #: ST GERARD
P.O. #:
Project name: 500018734

Quality Assurance Report
Maxxam Job Number: MA877972

QA/QC Batch	QC Type	Parameter	Date Analyzed yyyy/mm/dd	Value	Recovery	Units	QC Limits
1570400 HRE	MATRIX SPIKE	. Lead (Pb)	2008/07/25		100	%	80 - 120
	Spiked Blank	. Lead (Pb)	2008/07/25		101	%	85 - 115
	Method Blank	. Lead (Pb)	2008/07/25	ND, RDL=0.5		ug/L	
	RPD	. Lead (Pb)	2008/07/25	NC		%	25
1570417 HRE	MATRIX SPIKE	. Lead (Pb)	2008/07/25		100	%	80 - 120
	Spiked Blank	. Lead (Pb)	2008/07/25		97	%	85 - 115
	Method Blank	. Lead (Pb)	2008/07/25	ND, RDL=0.5		ug/L	
	RPD	. Lead (Pb)	2008/07/25	1.2		%	25

ND = Not detected
 NC = Non-calculable
 RPD = Relative Percent Difference
 SPIKE = Fortified sample

Validation Signature Page

Maxxam Job #: A877972

The analytical data and all QC contained in this report were reviewed and validated by the following individual(s).

Cristina Nervo

CHRISTINA NERVO, Scientific Services

=====

Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.



6740 Campobello Rd
Mississauga
Ontario L5N 2L8

REGULATION 243 DRINKING WATER SUB CHAIN OF CUSTODY RECORD

Tel: 905-817-5700
Fax: 905-817-5779
Toll Free: 1-800-563-6266

Samples are Regulated
Under Reg 243

17-Jul-08 14:46

JAMES ASPIN



A877972

FW ENV-756

CAM FCD-00196/2

Report to:

Company Name: Dufferin-Peel Catholic District School Board
Client Contact: Roland Lawton, General Manager of Physical Plant Facilities
Address: 5685 Keaton Cres., Mississauga ON L5R 3H5
E-Mail: roland.lawton@dpcdsb.org Tel: 905-890-0708 ext. 23204
Project: Regulation 243 Fax: 905-890-5841
Quote #: A83456 P.O.#: BN500636
Lab Service Rep.: James Aspin

Invoice to: Same as "Report to"

DW METALS
PRESERVATION
CHECKED

ANALYSIS REQUESTED	
Specify School Type	M.O.E. Assigned Water System Number
School	500018734
Private School	
Day nursery	

Sample Identification	Sample Type (Please Specify Standing / Flushed)	Date Sampled	Time Sampled	Sampler Initials	Field Chlorine		No. of Bottles	Lead
					Total	Free		
St. Gerard								
Kitchen - Staffroom	Standing	15-7-08	6:25 AM				1	X
Kitchen - Staffroom	Flushed	15-7-08	7:05 AM				1	X

SIF : Sample Inspection
Resolved By: _____
Date: _____

* In accordance with the Safe Drinking Water Act (2002), notification of adverse results must be reported. Failure to do so is an offence under the act. It is IMPERATIVE that the areas shaded in grey be filled out.

TAT (Turnaround Time)	Notification Information for School, Private Schools & Day Nurseries	LABORATORY USE ONLY
RUSH TAT MUST HAVE PRIOR APPROVAL	Facility Name: St. Gerard Facility Address: 1300 McBride Avenue Mississauga, ON L5C 1M8	Received By: <i>J. Aspin</i> Date: 08/07/08 Time: 14:46 Comments: 12/13/14°
<input checked="" type="checkbox"/> Normal (5-15 Working Days) <input type="checkbox"/> RUSH (Specify Below)	Tel #: _____ Fax #: _____ E-mail: _____	JUL 17 '08 14:46
Date Due:	Tel: 905-277-4512 Fax: 905-277-0647	
PN #		
Relinquished by (Sign Name): <i>John Lochane</i>	Date: 15-7-08	Laboratory Prepared Container: <input type="checkbox"/> Yes <input type="checkbox"/> No
Relinquished by (Print Name): <i>John Lochane</i>	Time: 7:05 AM	

NOTE: Samples submitted to Maxxam Analytics must be received below 10°C.
It is the Sampler's responsibility to ensure adequate cooling is provided so that the sample temperature is met.
Please ensure that a separate COC document is used for each individual sample.