

Your P.O. #: BN500116
Your Project #: ST JOHN OF THE CROSS
Site:500018747
Your C.O.C. #: na

Attention: Roland Lawton
Dufferin Peel Catholic District...
School Board
40 Matheson Blvd W
Mississauga, ON
L5R 1C5

Report Date: 2008/07/25

CERTIFICATE OF ANALYSIS

MAXXAM JOB #: A878397
Received: 2008/07/17, 14:09

Sample Matrix: Water
Samples Received: 2

Analyses	Quantity	Date Extracted	Date Analyzed	Laboratory Method	Method Reference
Metals Analysis by ICPMS (as received) ①	1	2008/07/24	2008/07/24	CAM SOP-00447	EPA 6020
Metals Analysis by ICPMS (as received) ①	1	2008/07/25	2008/07/25	CAM SOP-00447	EPA 6020

* RPDs calculated using raw data. The rounding of final results may result in the apparent difference.

(1) Metals analysis was performed on the sample 'as received'.

Encryption Key

Please direct all questions regarding this Certificate of Analysis to your Project Manager.

JAMES ASPIN, Project Manager
Email: James.Aspin@maxxamAnalytics.com
Phone# (905) 817-5700 Ext:5771

=====
Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.

For Service Group specific validation please refer to the Validation Signature Page

Total cover pages: 1

Maxxam Job #: A878397
Report Date: 2008/07/25

Dufferin Peel Catholic District...
Client Project #: ST JOHN OF THE CROSS
Project name: 500018747
Your P.O. #: BN500116
Sampler Initials: AV

ELEMENTS BY ATOMIC SPECTROSCOPY (WATER)

Maxxam ID			Z97017		Z97018		
Sampling Date			2008/07/17 07:23		2008/07/17 08:04		
COC Number			na		na		
	Units	Criteria	ST.JOHN OF THE CROSS KITCHEN- STAFFROOM STANDING	QC Batch	ST.JOHN OF THE CROSS KITCHEN- STAFFROOM FLUSHED	RDL	QC Batch

Metals							
. Lead (Pb)	mg/L	0.01	0.0067	1569253	0.0024	0.0005	1570400

RDL = Reportable Detection Limit
QC Batch = Quality Control Batch
CRITERIA: Ontario Regulation 243/07 made under the Safe Drinking Water Act, 2002.

Released June 7, 2007

Maxxam Job #: A878397
Report Date: 2008/07/25

Dufferin Peel Catholic District...
Client Project #: ST JOHN OF THE CROSS
Project name: 500018747
Your P.O. #: BN500116
Sampler Initials: AV

Package 1	8.7°C
-----------	-------

Each temperature is the average of up to three cooler temperatures taken at receipt

GENERAL COMMENTS

Results relate only to the items tested.

Dufferin Peel Catholic District...
Attention: Roland Lawton
Client Project #: ST JOHN OF THE CROSS
P.O. #: BN500116
Project name: 500018747

Quality Assurance Report
Maxxam Job Number: MA878397

QA/QC Batch	QC Type	Parameter	Date Analyzed yyyy/mm/dd	Value	Recovery	Units	QC Limits
1569253 JBW	MATRIX SPIKE	. Lead (Pb)	2008/07/24		106	%	80 - 120
	Spiked Blank	. Lead (Pb)	2008/07/24		99	%	85 - 115
	Method Blank	. Lead (Pb)	2008/07/24	ND, RDL=0.5		ug/L	
	RPD	. Lead (Pb)	2008/07/24	NC		%	25
1570400 HRE	MATRIX SPIKE	. Lead (Pb)	2008/07/25		100	%	80 - 120
	Spiked Blank	. Lead (Pb)	2008/07/25		101	%	85 - 115
	Method Blank	. Lead (Pb)	2008/07/25	ND, RDL=0.5		ug/L	
	RPD	. Lead (Pb)	2008/07/25	NC		%	25

ND = Not detected
 NC = Non-calculable
 RPD = Relative Percent Difference
 SPIKE = Fortified sample

Validation Signature Page

Maxxam Job #: A878397

The analytical data and all QC contained in this report were reviewed and validated by the following individual(s).



TROY CARRIERE, B.Sc., C.Chem, Scientific Specialist

=====

Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.

**DV METALS
PRESERVATION
CHECKED**

6740 Campobello Rd
Mississauga
Ontario L5N 2L8

17-Jul-08 14:09

JAMES ASPIN



A878397

NNA

ENV-998

Fax: 905-817-5779
Toll Free: 1-800-563-6266

**DRINKING WATER
BODY RECORD**
Samples are Regulated
Under Reg 243

SIF : Sample Inspection
Resolved By: _____
Date: _____

Maxxam Analytics Inc.
CAM FCD-00196/2

Report to:

Company Name: Dufferin-Peel Catholic District School Board
Client Contact: Roland Lawton, General Manager of Physical Plant Facilities
Address: 5685 Keaton Cres., Mississauga ON L5R 3H5
E-Mail: roland.lawton@dpcdsb.org Tel: 905-890-0708 ext. 23204
Project: Regulation 243 Fax: 905-890-5841
Quote #: A83456 P.O.#: BN500636
Lab Service Rep.: James Aspin

Invoice to: Same as "Report to"

PAGE _____ OF _____	
ANALYSIS REQUESTED	
Specify School Type	M.O.E. Assigned Water System Number 500018747
School	
Private School	
Day nursery	

Sample Identification	Sample Type (Please Specify Standing / Flushed)	Date Sampled	Time Sampled	Sampler Initials	Field Chlorine		No. of Bottles	Lead	School	Private School	Day nursery
					Total	Free					
St. John of the Cross Kitchen - Staffroom	Standing	17-7-08	7:23	ADV.			1	x	x		
Kitchen - Staffroom	Flushed	17-7-08	8:04	A.DV.			1	x	x		

* In accordance with the Safe Drinking Water Act (2002), notification of adverse results must be reported to the SAC, MOH & Waterworks.
Failure to do so is an offence under the act. It is **IMPERATIVE** that the areas shaded in grey be filled out completely before submitting to Maxxam Analytics.

TAT (Turnaround Time)	Notification Information for School, Private Schools & Day Nurseries		LABORATORY USE ONLY	
RUSH TAT MUST HAVE PRIOR APPROVAL	Facility Name: St. John of The Cross	Medical Officer of Health	Received By: <i>JK Jiang Wang</i>	
<input checked="" type="checkbox"/> Normal (5-15 Working Days)	Facility Address: 3180 Aquitaine Avenue	Region: Peel Region	Date: <i>08/07/08</i>	Time: <i>14:09</i>
<input type="checkbox"/> RUSH (Specify Below)	Mississauga, ON L5N 3S5	Tel #:	Comments: <i>81919</i>	
Date Due:	Tel: 905-824-3058 Fax: 905-824-4413	Fax #:		
PN #	E-mail:	Email:		
Relinquished by (Sign Name): <i>A. Del Vecchio</i>	Date: <i>17-JULY-08</i>	Laboratory Prepared Container:		
Relinquished by (Print Name): <i>ANGELO DEL VECCHIO</i>	Time: <i>8:04</i>	<input type="checkbox"/> Yes <input type="checkbox"/> No		

NOTE: Samples submitted to Maxxam Analytics must be received below 10°C.
It is the Sampler's responsibility to ensure adequate cooling is provided so that this receiving temperature is met.
Please ensure that a separate COC document is used for each indiv