

Your P.O. #: BN500636  
Your Project #: ST MARGUERITE BOURGEOYS  
Site:500012351  
Your C.O.C. #: NA

**Attention: Roland Lawton**  
Dufferin Peel Catholic District...  
School Board  
40 Matheson Blvd W  
Mississauga, ON  
L5R 1C5

**Report Date: 2008/08/07**

**CERTIFICATE OF ANALYSIS**

**MAXXAM JOB #: A885300**  
**Received: 2008/08/01, 13:17**

Sample Matrix: Water  
# Samples Received: 2

Analyses	Quantity	Date Extracted	Date Analyzed	Laboratory Method	Method Reference
Metals Analysis by ICPMS (as received) 0	2	2008/08/06	2008/08/07	CAM SOP-00447	EPA 6020

\* RPDs calculated using raw data. The rounding of final results may result in the apparent difference.

(1) Metals analysis was performed on the sample 'as received'.

**Encryption Key**

Please direct all questions regarding this Certificate of Analysis to your Project Manager.

JAMES ASPIN, Project Manager  
Email: James.Aspin@maxxamAnalytics.com  
Phone# (905) 817-5700 Ext:5771

=====  
Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.

For Service Group specific validation please refer to the Validation Signature Page

Total cover pages: 1

Maxxam Job #: A885300  
Report Date: 2008/08/07

Dufferin Peel Catholic District...  
Client Project #: ST MARGUERITE BOURGEOYS  
Project name: 500012351  
Your P.O. #: BN500636

**ELEMENTS BY ATOMIC SPECTROSCOPY (WATER)**

Maxxam ID			AC8117	AC8118		
Sampling Date			2008/07/31 07:10	2008/07/31 07:46		
COC Number			NA	NA		
	<b>Units</b>	<b>Criteria</b>	<b>ST MARGUERITE BOURGEOYS STANDING</b>	<b>ST MARGUERITE BOURGEOYS FLUSHED</b>	<b>RDL</b>	<b>QC Batch</b>

<b>Metals</b>						
. Lead (Pb)	mg/L	<b>0.01</b>	0.0040	0.0015	0.0005	1579769

RDL = Reportable Detection Limit  
QC Batch = Quality Control Batch  
CRITERIA: Ontario Regulation 243/07 made under the Safe Drinking Water Act, 2002.  
Released June 7, 2007

Maxxam Job #: A885300  
Report Date: 2008/08/07

Dufferin Peel Catholic District...  
Client Project #: ST MARGUERITE BOURGEOYS  
Project name: 500012351  
Your P.O. #: BN500636

Package 1	8.7°C
-----------	-------

Each temperature is the average of up to three cooler temperatures taken at receipt

**GENERAL COMMENTS**

**Results relate only to the items tested.**

Dufferin Peel Catholic District...  
Attention: Roland Lawton  
Client Project #: ST MARGUERITE BOURGEOYS  
P.O. #: BN500636  
Project name: 500012351

Quality Assurance Report  
Maxxam Job Number: MA885300

QA/QC Batch	QC Type	Parameter	Date Analyzed yyyy/mm/dd	Value	Recovery	Units	QC Limits
1579769 HRE	MATRIX SPIKE	. Lead (Pb)	2008/08/07		98	%	80 - 120
	Spiked Blank	. Lead (Pb)	2008/08/07		98	%	85 - 115
	Method Blank	. Lead (Pb)	2008/08/07	ND, RDL=0.0005		mg/L	
	RPD	. Lead (Pb)	2008/08/07	0.2		%	25


ND = Not detected  
RPD = Relative Percent Difference  
SPIKE = Fortified sample

**Validation Signature Page**

**Maxxam Job #: A885300**

---

The analytical data and all QC contained in this report were reviewed and validated by the following individual(s).



---

BRAD NEWMAN, Scientific Specialist

=====

Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.



6740 Campobello Rd  
Mississauga  
Ontario L5N 2L8

REGULATION 243 DRINKING WATER SUBM  
**CHAIN OF CUSTODY RECORD**

Tel: 905-817-5700  
Fax: 905-817-5779  
Toll Free: 1-800-563-6266

Samples are Regulated  
Under Reg 243

1-Aug-08 13:17  
JAMES ASPIN  
 A885300  
M\_O ENV-813

**Report to:**

Company Name: Dufferin-Peel Catholic District School Board  
Client Contact: Roland Lawton, General Manager of Physical Plant Facilities  
Address: 5685 Keaton Cres., Mississauga ON L5R 3H5  
E-Mail: roland.lawton@dpcdsb.org Tel: 905-890-0708 ext. 23204  
Project: Regulation 243 Fax: 905-890-5841  
Quote #: A83456 P.O.#: BN500636  
Lab Service Rep.: James Aspen

Invoice to:  Same as "Report to"

PAGE \_\_\_\_ OF \_\_\_\_  
**ANALYSIS REQUESTED**  
Specify School Type  
M.O.E. Assigned Water System Number  
500012351  
School Private School Day nursery

Sample Identification	Sample Type (Please Specify Standing / Flushed)	Date Sampled	Time Sampled	Sampler Initials	Field Chlorine		No. of Bottles	Lead	School	Private School	Day nursery	M.O.E. Assigned Water System Number
					Total	Free						
St. Marguerite Bourgeoys												
Kitchen - Staffroom	Standing	07/31	7:10 AM				1	X	X			500012351
Kitchen - Staffroom	Flushed	07/31	7:46 AM				1	X	X			

DW METALS PRESERVATION CHECKED

\* In accordance with the Safe Drinking Water Act (2002), notification of adverse results must be re... Failure to do so is an offence under the act. It is IMPERATIVE that the areas shaded in grey be... Analytics.

TAT (Turnaround Time)	Notification Information for School, Private Schools & Daycares		LABORATORY USE ONLY
<b>RUSH TAT MUST HAVE PRIOR APPROVAL</b>	Facility Name: St. Marguerite Bourgeoys	Medical Officer of Health	Received By: ZOFIA LEATH
	Facility Address: 550 North Park Drive	Region: Peel Region	Date: 08/08/08 Time: 13:17
<input checked="" type="checkbox"/> Normal (5-15 Working Days)	Brampton, ON L6S 4J8		Comments:
<input type="checkbox"/> RUSH (Specify Below)		Tel #:	
Date Due:	Tel: 905-454-3979 Fax: 905-454-5591	Fax #:	
PN #	E-mail:	Email:	
Relinquished by (Sign Name): David Rizon	Date: July 31/08	Laboratory Prepared Container:	
Relinquished by (Print Name): DAVID MAKIAMPILLAI	Time: 7:46 AM	<input type="checkbox"/> Yes <input type="checkbox"/> No	

SIF: Sample Inspection  
Resolved By: \_\_\_\_\_  
Date: \_\_\_\_\_

...submitted to Maxxam Analytics must be received below 10°C. Ensure adequate cooling is provided so that this receiving temperature is met. that a separate COC document is used for each indiv

'08 AUG 1 13:17  
8/8/008