

Your Project #: ST. MARIA GORETTI
Site:500012364
Your C.O.C. #: na

Attention: Roland Lawton
Dufferin Peel Catholic District...
School Board
40 Matheson Blvd W
Mississauga, ON
L5R 1C5

Report Date: 2008/08/12

CERTIFICATE OF ANALYSIS

MAXXAM JOB #: A885775
Received: 2008/07/17, 14:38

Sample Matrix: Water
Samples Received: 2

Analyses	Quantity	Date Extracted	Date Analyzed	Laboratory Method	Method Reference
Metals Analysis by ICPMS (as received) ①	1	2008/08/07	2008/08/07	CAM SOP-00447	EPA 6020
Metals Analysis by ICPMS (as received) ①	1	2008/08/11	2008/08/11	CAM SOP-00447	EPA 6020

* RPDs calculated using raw data. The rounding of final results may result in the apparent difference.

(1) Metals analysis was performed on the sample 'as received'.

Encryption Key

Please direct all questions regarding this Certificate of Analysis to your Project Manager.

JAMES ASPIN, Project Manager
Email: James.Aspin@maxxamAnalytics.com
Phone# (905) 817-5700 Ext:5771

=====
Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.

For Service Group specific validation please refer to the Validation Signature Page

Total cover pages: 1

Maxxam Job #: A885775
Report Date: 2008/08/12

Dufferin Peel Catholic District...
Client Project #: ST. MARIA GORETTI
Project name: 500012364

ELEMENTS BY ATOMIC SPECTROSCOPY (WATER)

Maxxam ID			AD0425		AD0426		
Sampling Date			2008/07/15 06:30		2008/07/15 07:10		
COC Number			na		na		
	Units	Criteria	ST. MARIA GORETTI KITCHEN STAFFROOM STANDING	QC Batch	ST. MARIA GORETTI KITCHEN STAFFROOM FLUSHED	RDL	QC Batch

Metals							
. Lead (Pb)	mg/L	0.01	ND	1580482	ND	0.0005	1583093

ND = Not detected
RDL = Reportable Detection Limit
QC Batch = Quality Control Batch
CRITERIA: Ontario Regulation 243/07 made under the Safe Drinking Water Act, 2002.

Released June 7, 2007

Maxxam Job #: A885775
Report Date: 2008/08/12

Dufferin Peel Catholic District...
Client Project #: ST. MARIA GORETTI
Project name: 500012364

Package 1	14.7°C
-----------	--------

Each temperature is the average of up to three cooler temperatures taken at receipt

GENERAL COMMENTS

Results relate only to the items tested.

Dufferin Peel Catholic District...
Attention: Roland Lawton
Client Project #: ST. MARIA GORETTI
P.O. #:
Project name: 500012364

Quality Assurance Report
Maxxam Job Number: MA885775

QA/QC Batch	QC Type	Parameter	Date Analyzed yyyy/mm/dd	Value	Recovery	Units	QC Limits
1580482 HRE	MATRIX SPIKE	. Lead (Pb)	2008/08/07		99	%	80 - 120
	Spiked Blank	. Lead (Pb)	2008/08/07		101	%	85 - 115
	Method Blank	. Lead (Pb)	2008/08/07	ND, RDL=0.0005		mg/L	
	RPD	. Lead (Pb)	2008/08/07	NC		%	25
1583093 HRE	MATRIX SPIKE	. Lead (Pb)	2008/08/11		107	%	80 - 120
	Spiked Blank	. Lead (Pb)	2008/08/11		106	%	85 - 115
	Method Blank	. Lead (Pb)	2008/08/11	ND, RDL=0.0005		mg/L	
	RPD	. Lead (Pb)	2008/08/11	NC		%	25

ND = Not detected
NC = Non-calculable
RPD = Relative Percent Difference
SPIKE = Fortified sample

Validation Signature Page

Maxxam Job #: A885775

The analytical data and all QC contained in this report were reviewed and validated by the following individual(s).

Cristina Nervo

CHRISTINA NERVO, Scientific Services

=====

Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.

Ma

17-Jul-08 14:38

JAMES ASPIN



A885775

NNA

ENV-667

REGULATION 243 DRINKING WATER SUBMISSION CHAIN OF CUSTODY RECORD

Tel: 905-817-5700
Fax: 905-817-5779
Toll Free: 1-800-563-6266

Samples are Regulated
Under Reg 243

Maxxam Analytics Inc.
CAM FCD-00196/2

Mississauga
Ontario L5N 2L8

Report to:

Company Name:	Dufferin-Peel Catholic District School Board		
Client Contact:	Roland Lawton, General Manager of Physical Plant Facilities		
Address:	5685 Keaton Cres., Mississauga ON L5R 3H5		
E-Mail:	roland.lawton@dpcdsb.org	Tel:	905-890-0708 ext. 23204
Project:	Regulation 243	Fax:	905-890-5841
Quote #:	A83456	P.O.#:	BN500636
Lab Service Rep.:	James Aspen		

Invoice to: Same as "Report to"

DW METALS
PRESERVATION
CHECKED

PAGE ____ OF ____

ANALYSIS REQUESTED

Specify School Type	M.O.E. Assigned Water System Number 500012364
School	
Private School	
Day nursery	

Sample Identification	Sample Type (Please Specify Standing / Flushed)	Date Sampled	Time Sampled	Sampler Initials	Field Chlorine		No. of Bottles	Lead
					Total	Free		
St. Maria Goretti Kitchen - Staffroom	Standing	7/15	6:30 AM				1	x
Kitchen - Staffroom	Flushed	7/15	7:10 AM				1	x

* In accordance with the Safe Drinking Water Act (2002), notification of adverse results must be reported to the SAC, MOH & Waterworks. Failure to do so is an offence under the act. It is IMPERATIVE that the areas shaded in grey be filled out completely before submitting to Maxxam Analytics.

TAT (Turnaround Time)	Notification Information for School, Private Schools & Day Nurseries		LABORATORY USE ONLY
RUSH TAT MUST HAVE PRIOR APPROVAL	Facility Name: St. Maria Goretti	Medical Officer of Health	Received By: <i>[Signature]</i>
<input checked="" type="checkbox"/> Normal (5-15 Working Days)	Facility Address: 121 Royal Orchard Drive	Region: Peel Region	Date: 08/07/14 Time: 14:38
<input type="checkbox"/> RUSH (Specify Below)	Brampton, ON L6X 4K9	Tel #:	Comments: 14/15/15
Date Due:	Tel: 905-454-4458 Fax: 905-454-0360	Fax #:	07/17/08 14:38
PN #	E-mail:	Email:	
Relinquished by (Sign Name): <i>Lydia Clatoff</i>	Date: <i>July 17/08</i>	Laboratory Prepared Container:	<input type="checkbox"/> Yes <input type="checkbox"/> No
Relinquished by (Print Name): <i>L. Clatoff</i>	Time: <i>2:55pm</i>		

NOTE: Samples submitted to Maxxam Analytics must be received below 10°C. It is the Sampler's responsibility to ensure adequate cooling is provided so that this receiving temperature is met. Please ensure that a separate COC document is used for each indiv