

Your P.O. #: BN500116
Your Project #: ST MARK
Site:500019189
Your C.O.C. #: na

Attention: Roland Lawton
Dufferin Peel Catholic District...
School Board
40 Matheson Blvd W
Mississauga, ON
L5R 1C5

Report Date: 2008/07/28

CERTIFICATE OF ANALYSIS

MAXXAM JOB #: A878418
Received: 2008/07/17, 14:48

Sample Matrix: Water
Samples Received: 2

Analyses	Quantity	Date Extracted	Date Analyzed	Laboratory Method	Method Reference
Metals Analysis by ICPMS (as received) 0	2	2008/07/25	2008/07/26	CAM SOP-00447	EPA 6020

* RPDs calculated using raw data. The rounding of final results may result in the apparent difference.

(1) Metals analysis was performed on the sample 'as received'.

Encryption Key

Please direct all questions regarding this Certificate of Analysis to your Project Manager.

JAMES ASPIN, Project Manager
Email: James.Aspin@maxxamAnalytics.com
Phone# (905) 817-5700 Ext:5771

=====

Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.

For Service Group specific validation please refer to the Validation Signature Page

Total cover pages: 1

Maxxam Job #: A878418
Report Date: 2008/07/28

Dufferin Peel Catholic District...
Client Project #: ST MARK
Project name: 500019189
Your P.O. #: BN500116
Sampler Initials: BH

ELEMENTS BY ATOMIC SPECTROSCOPY (WATER)

Maxxam ID			Z97114	Z97115		
Sampling Date			2008/07/17 06:50	2008/07/17 07:25		
COC Number			na	na		
	Units	Criteria	ST. MARK KITCHEN- STAFFROOM STANDING	ST. MARK KITCHEN- STAFFROOM FLUSHED	RDL	QC Batch

Metals						
. Lead (Pb)	mg/L	0.01	0.0022	0.0009	0.0005	1571011

RDL = Reportable Detection Limit
QC Batch = Quality Control Batch
CRITERIA: Ontario Regulation 243/07 made under the Safe Drinking Water Act, 2002.
Released June 7, 2007

Maxxam Job #: A878418
Report Date: 2008/07/28

Dufferin Peel Catholic District...
Client Project #: ST MARK
Project name: 500019189
Your P.O. #: BN500116
Sampler Initials: BH

Package 1	12.3°C
-----------	--------

Each temperature is the average of up to three cooler temperatures taken at receipt

GENERAL COMMENTS

Results relate only to the items tested.

Dufferin Peel Catholic District...
Attention: Roland Lawton
Client Project #: ST MARK
P.O. #: BN500116
Project name: 500019189

Quality Assurance Report
Maxxam Job Number: MA878418

QA/QC Batch	QC Type	Parameter	Date Analyzed yyyy/mm/dd	Value	Recovery	Units	QC Limits
1571011 HRE	MATRIX SPIKE	. Lead (Pb)	2008/07/26		100	%	80 - 120
	Spiked Blank	. Lead (Pb)	2008/07/26		100	%	85 - 115
	Method Blank	. Lead (Pb)	2008/07/26	ND, RDL=0.5		ug/L	
	RPD	. Lead (Pb)	2008/07/26	1.5		%	25

ND = Not detected
RPD = Relative Percent Difference
SPIKE = Fortified sample

Validation Signature Page

Maxxam Job #: A878418

The analytical data and all QC contained in this report were reviewed and validated by the following individual(s).

Cristina Nervo

CHRISTINA NERVO, Scientific Services

=====

Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.

SIF : Sample Inspection
 Resolved By: _____
 Date: _____

Mississauga
 Ontario L5N 2L8

17-Jul-08 14:48
 JAMES ASPIN
 A878418
 NNA ENV-998

RECEIVED DRINKING WATER SUPPLY RECORD
 C
 Tel: 905-817-5779
 Fax: 905-817-5779
 Toll Free: 1-800-563-6266

DW METALS
 PRESERVATION
 CHECKED

Maxxam Analytics Inc.
 CAM FCD-00196/2

Report to:

Company Name: Dufferin-Peel Catholic District School Board
 Client Contact: Roland Lawton, General Manager of Physical Plant Facilities
 Address: 5685 Keaton Cres., Mississauga ON L5R 3H5
 E-Mail: roland.lawton@dpcdsb.org Tel: 905-890-0708 ext. 23204
 Project: Regulation 243 Fax: 905-890-5841
 Quote #: A83456 P.O.#: BN500636
 Lab Service Rep.: James Aspen

Invoice to: Same as "Report to"

PAGE ____ OF ____

ANALYSIS REQUESTED			
Specify School Type	Private School	Day nursery	M.O.E. Assigned Water System Number
			500019189

Sample Identification	Sample Type (Please Specify Standing / Flushed)	Date Sampled	Time Sampled	Sampler Initials	Field Chlorine		No. of Bottles	Lead	School	Private School	Day nursery
					Total	Free					
St. Mark											
Kitchen - Staffroom	Standing	07-07-08	6:50				1	x	x		
Kitchen - Staffroom	Flushed	07-07-08	7:25				1	x	x		

* In accordance with the Safe Drinking Water Act (2002), notification of adverse results must be reported to the SAC, MOH & Waterworks.
 Failure to do so is an offence under the act. It is **IMPERATIVE** that the areas shaded in grey be filled out completely before submitting to Maxxam Analytics.

TAT (Turnaround Time)	Notification Information for School, Private Schools & Day Nurseries			LABORATORY USE ONLY
RUSH TAT MUST HAVE PRIOR APPROVAL	Facility Name: St. Mark	Medical Officer of Health		Received By: <i>J. J. [Signature]</i>
<input checked="" type="checkbox"/> Normal (5-15 Working Days)	Facility Address: 3675 Sawmill Valley Drive	Region: Peel Region		Date: 08/07/08 Time: 14:48
<input type="checkbox"/> RUSH (Specify Below)	Mississauga, ON L5L 2Z5	Tel #:		Comments: 12/12/13
Date Due:	Tel: 905-820-9477 Fax: 905-820-1229	Fax #:		JUL 17 '08 14:48
PN #	E-mail:	Email:		
Relinquished by (Sign Name): <i>B. Herczeg</i>	Date: <i>Jul 17, 08</i>	Laboratory Prepared Container:		
Relinquished by (Print Name): <i>BARB HERCEG</i>	Time: <i>7:25 AM</i>	<input type="checkbox"/> Yes <input type="checkbox"/> No		

NOTE: Samples submitted to Maxxam Analytics must be received below 10°C.
 It is the Sampler's responsibility to ensure adequate cooling is provided so that this receiving temperature is met.
 Please ensure that a separate COC document is used for each indiv