

Your P.O. #: BN500636  
Your Project #: ST PAUL SS R243  
Site:500018812  
Your C.O.C. #: NA

**Attention: Roland Lawton**  
Dufferin Peel Catholic District...  
School Board  
40 Matheson Blvd W  
Mississauga, ON  
L5R 1C5

**Report Date: 2008/07/24**

**CERTIFICATE OF ANALYSIS**

**MAXXAM JOB #: A878053**  
**Received: 2008/07/17, 13:29**

Sample Matrix: Water  
# Samples Received: 2

Analyses	Quantity	Date Extracted	Date Analyzed	Laboratory Method	Method Reference
Metals Analysis by ICPMS (as received) 0	2	2008/07/22	2008/07/24	CAM SOP-00447	EPA 6020

\* RPDs calculated using raw data. The rounding of final results may result in the apparent difference.

(1) Metals analysis was performed on the sample 'as received'.

**Encryption Key**

Please direct all questions regarding this Certificate of Analysis to your Project Manager.

JAMES ASPIN, Project Manager  
Email: James.Aspin@maxxamAnalytics.com  
Phone# (905) 817-5700 Ext:5771

=====  
Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.

For Service Group specific validation please refer to the Validation Signature Page

Total cover pages: 1

Maxxam Job #: A878053  
Report Date: 2008/07/24

Dufferin Peel Catholic District...  
Client Project #: ST PAUL SS R243  
Project name: 500018812  
Your P.O. #: BN500636

**ELEMENTS BY ATOMIC SPECTROSCOPY (WATER)**

Maxxam ID			Z95672	Z95673		
Sampling Date			2008/07/17 06:47	2008/07/17 07:25		
COC Number			NA	NA		
	<b>Units</b>	<b>Criteria</b>	<b>ST. PAUL S.S. KITCHEN-STAFFROOM/STANDING</b>	<b>ST. PAUL S.S. KITCHEN-STAFFROOM/FLUSHED</b>	<b>RDL</b>	<b>QC Batch</b>

<b>Metals</b>						
Lead (Pb)	mg/L	<b>0.01</b>	0.0038	0.0011	0.0005	1567574

RDL = Reportable Detection Limit  
QC Batch = Quality Control Batch  
CRITERIA: Ontario Regulation 243/07 made under the Safe Drinking Water Act, 2002.

Released June 7, 2007

Maxxam Job #: A878053  
Report Date: 2008/07/24

Dufferin Peel Catholic District...  
Client Project #: ST PAUL SS R243  
Project name: 500018812  
Your P.O. #: BN500636

Package 1	7.3°C
-----------	-------

Each temperature is the average of up to three cooler temperatures taken at receipt

**GENERAL COMMENTS**

**Results relate only to the items tested.**

Dufferin Peel Catholic District...  
Attention: Roland Lawton  
Client Project #: ST PAUL SS R243  
P.O. #: BN500636  
Project name: 500018812

Quality Assurance Report  
Maxxam Job Number: MA878053

QA/QC Batch	QC Type	Parameter	Date Analyzed yyyy/mm/dd	Value	Recovery	Units	QC Limits
1567574 JBW	MATRIX SPIKE	. Lead (Pb)	2008/07/24		104	%	80 - 120
	Spiked Blank	. Lead (Pb)	2008/07/24		99	%	85 - 115
	Method Blank	. Lead (Pb)	2008/07/24	ND, RDL=0.5		ug/L	
	RPD	. Lead (Pb)	2008/07/24	0.8		%	25

ND = Not detected  
RPD = Relative Percent Difference  
SPIKE = Fortified sample

**Validation Signature Page**

**Maxxam Job #: A878053**

---

The analytical data and all QC contained in this report were reviewed and validated by the following individual(s).

*Cristina Nervo*

---

CHRISTINA NERVO, Scientific Services

=====

Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.



6740 Campobello Rd  
Mississauga  
Ontario L5N 2L8

# REGULATION 243 DRINKING WATER SUBMIS CHAIN OF CUSTODY RECORD

Tel: 905-817-5700  
Fax: 905-817-5779  
Toll Free: 1-800-563-6266

Samples are Regulated  
Under Reg 243

17-Jul-08 13:29  
JAMES ASPIN  
 A878053  
FW ENV-925  
CAM FCD-00196/2

**Report to:**

Company Name: Dufferin-Peel Catholic District School Board  
Client Contact: Roland Lawton, General Manager of Physical Plant Facilities  
Address: 5685 Keaton Cres., Mississauga ON L5R 3H5  
E-Mail: [roland.lawton@dpcdsb.org](mailto:roland.lawton@dpcdsb.org) Tel: 905-890-0708 ext. 23204  
Project: Regulation 243 Fax: 905-890-5841  
Quote #: A83456 P.O.#: BN500636  
Lab Service Rep.: James Aspen

Invoice to:  Same as "Report to"

**DW METALS  
PRESERVATION  
CHECKED**

ANALYSIS REQUESTED	
Specify School Type	M.O.E. Assigned Water System Number
School	500018812
Private School	
Day nursery	

Sample Identification	Sample Type (Please Specify Standing / Flushed)	Date Sampled	Time Sampled	Sampler Initials	Field Chlorine		No. of Bottles	Lead
					Total	Free		
St. Paul S.S.								
Kitchen - Staffroom	Standing	07/17/08	6:47	TD			1	X
Kitchen - Staffroom	Flushed	07/17/08	7:25 AM	TD			1	X

In accordance with the Safe Drinking Water Act (2002), notification of adverse results must be reported to the SAC, MOH & Waterworks. Failure to do so is an offence under the act. It is **IMPERATIVE** that the areas shaded in grey be filled out completely before submitting to Maxxam Analytics.

TAT (Turnaround Time)	Notification Information for School, Private Schools & Day Nurseries	LABORATORY USE ONLY
<b>RUSH TAT MUST HAVE PRIOR APPROVAL</b>	Facility Name: St. Paul S.S. Facility Address: 815 Atwater Avenue Mississauga, ON L5E 1L8	Received By: <i>ZOFIA W...</i> Date: 08/07/08 Time: 13:29
<input checked="" type="checkbox"/> Normal (5-15 Working Days)		Comments:
<input type="checkbox"/> RUSH (Specify Below)		
Date Due:	Tel: 905-278-3994 Fax: 905-278-1163	
Lab #	E-mail:	
Relinquished by (Sign Name): <i>Jay DiRenzo</i>	Date: JULY 17/2008	Laboratory Prepared Container:
Relinquished by (Print Name): TONY DIRENZO	Time: 7:30 AM	<input type="checkbox"/> Yes <input type="checkbox"/> No

**NOTE:** Samples submitted to Maxxam Analytics must be received below 10°C. It is the Sampler's responsibility to ensure adequate cooling is provided so that this receiving temperature is met. Please ensure that a separate COC document is used for each indiv

08 JUL 17 13:29  
7/17/08