

Your P.O. #: BN500636
Your Project #: ST RICHARD
Site:500019527
Your C.O.C. #: na

Attention: Roland Lawton
Dufferin Peel Catholic District...
School Board
40 Matheson Blvd W
Mississauga, ON
L5R 1C5

Report Date: 2008/08/07

CERTIFICATE OF ANALYSIS

MAXXAM JOB #: A885141
Received: 2008/08/01, 13:07

Sample Matrix: Water
Samples Received: 2

Analyses	Quantity	Date Extracted	Date Analyzed	Laboratory Method	Method Reference
Metals Analysis by ICPMS (as received) ¶	2	2008/08/06	2008/08/07	CAM SOP-00447	EPA 6020

* RPDs calculated using raw data. The rounding of final results may result in the apparent difference.

(1) Metals analysis was performed on the sample 'as received'.

Encryption Key

Please direct all questions regarding this Certificate of Analysis to your Project Manager.

JAMES ASPIN, Project Manager
Email: James.Aspin@maxxamAnalytics.com
Phone# (905) 817-5700 Ext:5771

=====

Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.

For Service Group specific validation please refer to the Validation Signature Page

Total cover pages: 1

Maxxam Job #: A885141
Report Date: 2008/08/07

Dufferin Peel Catholic District...
Client Project #: ST RICHARD
Project name: 500019527
Your P.O. #: BN500636

ELEMENTS BY ATOMIC SPECTROSCOPY (WATER)

Maxxam ID			AC7088		
Sampling Date			2008/07/31 06:55		
COC Number			na		
	Units	Criteria	ST. RICHARD-KITCHEN-STAFFROOM-ST ANDING	RDL	QC Batch

Metals					
. Lead (Pb)	mg/L	0.01	ND	0.0005	1579769
<p>ND = Not detected RDL = Reportable Detection Limit QC Batch = Quality Control Batch CRITERIA: Ontario Regulation 243/07 made under the Safe Drinking Water Act, 2002. Released June 7, 2007</p>					

Maxxam ID			AC7089		
Sampling Date			2008/07/31 07:30		
COC Number			na		
	Units	Criteria	ST. RICHARD-KITCHEN-STAFFROOM-FL USHED	RDL	QC Batch

Metals					
. Lead (Pb)	mg/L	0.01	ND	0.0005	1579769
<p>ND = Not detected RDL = Reportable Detection Limit QC Batch = Quality Control Batch CRITERIA: Ontario Regulation 243/07 made under the Safe Drinking Water Act, 2002. Released June 7, 2007</p>					

Maxxam Job #: A885141
Report Date: 2008/08/07

Dufferin Peel Catholic District...
Client Project #: ST RICHARD
Project name: 500019527
Your P.O. #: BN500636

Package 1	9.3°C
-----------	-------

Each temperature is the average of up to three cooler temperatures taken at receipt

GENERAL COMMENTS

Results relate only to the items tested.

Dufferin Peel Catholic District...
Attention: Roland Lawton
Client Project #: ST RICHARD
P.O. #: BN500636
Project name: 500019527

Quality Assurance Report
Maxxam Job Number: MA885141


QA/QC Batch	QC Type	Parameter	Date Analyzed yyyy/mm/dd	Value	Recovery	Units	QC Limits
1579769 HRE	MATRIX SPIKE	. Lead (Pb)	2008/08/07		98	%	80 - 120
	Spiked Blank	. Lead (Pb)	2008/08/07		98	%	85 - 115
	Method Blank	. Lead (Pb)	2008/08/07	ND, RDL=0.0005		mg/L	
	RPD	. Lead (Pb)	2008/08/07	0.2		%	25

ND = Not detected
RPD = Relative Percent Difference
SPIKE = Fortified sample

Validation Signature Page

Maxxam Job #: A885141

The analytical data and all QC contained in this report were reviewed and validated by the following individual(s).



BRAD NEWMAN, Scientific Specialist

=====

Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.



6740 Campobello Rd
Mississauga
Ontario L5N 2L8

REGULATION 243 DRINKING WATER SUBMISSION
CHAIN OF CUSTODY RECORD

Tel: 905-817-5700
Fax: 905-817-5779
Toll Free: 1-800-563-6266

Samples are Regulated
Under Reg 243

1-Aug-08 13:07

JAMES ASPIN



A885141

CCO

ENV-813

Report to:

Company Name: Dufferin-Peel Catholic District School Board
Client Contact: Roland Lawton, General Manager of Physical Plant Facilities
Address: 5685 Keaton Cres., Mississauga ON L5R 3H5
E-Mail: roland.lawton@dpcdsb.org Tel: 905-890-0708 ext. 23204
Project: Regulation 243 Fax: 905-890-5841
Quote #: A83456 P.O.#: BN500636
Lab Service Rep.: James Aspin

Invoice to: Same as "Report to"

DW METALS
PRESERVATION
CHECKED

PAGE ____ OF ____

ANALYSIS REQUESTED

Specify School Type

M.O.E. Assigned Water System Number

Sample Identification	Sample Type (Please Specify Standing / Flushed)	Date Sampled	Time Sampled	Sampler Initials	Field Chlorine		No. of Bottles	Lead	Specify School Type			M.O.E. Assigned Water System Number
					Total	Free			School	Private School	Day nursery	
St. Richard												500019527
Kitchen - Staffroom	Standing	July 31/08	6:55	P.J			1	X	X			
Kitchen - Staffroom	Flushed	July 31/08	7:30	P.J			1	X	X			

* In accordance with the Safe Drinking Water Act (2002), notification of adverse results must be reported to the SAC, MOH & Waterworks. Failure to do so is an offence under the act. It is **IMPERATIVE** that the areas shaded in grey be filled out completely before submitting to Maxxam Analytics.

TAT (Turnaround Time)	Notification Information for School, Private Schools & Day Nurseries		LABORATORY USE ONLY
RUSH TAT MUST HAVE PRIOR APPROVAL	Facility Name: St. Richard	Medical Officer of Health	Received By: ZOFIA ZERETA
<input checked="" type="checkbox"/> Normal (5-15 Working Days)	Facility Address: 7270 Copenhagen Road	Region: Peel Region	Date: 08/08/08 Time: 13:07
<input type="checkbox"/> RUSH (Specify Below)	Mississauga, ON L5N 2C3	Tel #:	Comments:
Date Due:	Tel: 905-826-5572 Fax: 905-826-9554	Fax #:	
PN #	E-mail:	Email:	
Relinquished by (Sign Name): <i>[Signature]</i>	Date: July 31/08	Laboratory Prepared Container:	
Relinquished by (Print Name): STEVE PECZKOWSKI	Time: 7:55 AM	<input type="checkbox"/> Yes <input type="checkbox"/> No	

NOTE: Samples submitted to Maxxam Analytics must be...
It is the Sampler's responsibility to ensure adequate cooling is provided so...
Please ensure that a separate COC document is use

SIF : Sample Inspection

Page 6 of 6
Reserved By: _____

08 AUG 1 13:07
7/9/08