

Your P.O. #: BN500116
Your Project #: ST ROSE OF LIMA
Site:500013820
Your C.O.C. #: na

Attention: Roland Lawton
Dufferin Peel Catholic District...
School Board
40 Matheson Blvd W
Mississauga, ON
L5R 1C5

Report Date: 2008/07/28

CERTIFICATE OF ANALYSIS

MAXXAM JOB #: A878412
Received: 2008/07/17, 14:50

Sample Matrix: Water
Samples Received: 2

Analyses	Quantity	Date Extracted	Date Analyzed	Laboratory Method	Method Reference
Metals Analysis by ICPMS (as received) ①	1	2008/07/25	2008/07/25	CAM SOP-00447	EPA 6020
Metals Analysis by ICPMS (as received) ①	1	2008/07/25	2008/07/26	CAM SOP-00447	EPA 6020

* RPDs calculated using raw data. The rounding of final results may result in the apparent difference.

(1) Metals analysis was performed on the sample 'as received'.

Encryption Key

Please direct all questions regarding this Certificate of Analysis to your Project Manager.

JAMES ASPIN, Project Manager
Email: James.Aspin@maxxamAnalytics.com
Phone# (905) 817-5700 Ext:5771

=====
Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.

For Service Group specific validation please refer to the Validation Signature Page

Total cover pages: 1

Maxxam Job #: A878412
Report Date: 2008/07/28

Dufferin Peel Catholic District...
Client Project #: ST ROSE OF LIMA
Project name: 500013820
Your P.O. #: BN500116
Sampler Initials: AK

ELEMENTS BY ATOMIC SPECTROSCOPY (WATER)

Maxxam ID			Z97087		Z97088		
Sampling Date			2008/07/17 06:55		2008/07/17 07:32		
COC Number			na		na		
	Units	Criteria	ST. ROSE OF LIMA KITCHEN - STAFFROOM STANDING	QC Batch	ST. ROSE OF LIMA KITCHEN - STAFFROOM FLUSHED	RDL	QC Batch

Metals							
. Lead (Pb)	mg/L	0.01	0.0036	1570417	0.0018	0.0005	1570546

RDL = Reportable Detection Limit
QC Batch = Quality Control Batch
CRITERIA: Ontario Regulation 243/07 made under the Safe Drinking Water Act, 2002.
Released June 7, 2007

Maxxam Job #: A878412
Report Date: 2008/07/28

Dufferin Peel Catholic District...
Client Project #: ST ROSE OF LIMA
Project name: 500013820
Your P.O. #: BN500116
Sampler Initials: AK

Package 1	12.3°C
-----------	--------

Each temperature is the average of up to three cooler temperatures taken at receipt

GENERAL COMMENTS

Results relate only to the items tested.

Dufferin Peel Catholic District...
Attention: Roland Lawton
Client Project #: ST ROSE OF LIMA
P.O. #: BN500116
Project name: 500013820

Quality Assurance Report
Maxxam Job Number: MA878412

QA/QC Batch	QC Type	Parameter	Date Analyzed yyyy/mm/dd	Value	Recovery	Units	QC Limits
1570417 HRE	MATRIX SPIKE	. Lead (Pb)	2008/07/25		100	%	80 - 120
	Spiked Blank	. Lead (Pb)	2008/07/25		97	%	85 - 115
	Method Blank	. Lead (Pb)	2008/07/25	ND, RDL=0.5		ug/L	
	RPD	. Lead (Pb)	2008/07/25	1.2		%	25
1570546 JBW	MATRIX SPIKE [Z97088-01]	. Lead (Pb)	2008/07/26		107	%	80 - 120
	Spiked Blank	. Lead (Pb)	2008/07/26		97	%	85 - 115
	Method Blank	. Lead (Pb)	2008/07/26	ND, RDL=0.5		ug/L	
	RPD [Z97088-01]	. Lead (Pb)	2008/07/26	NC		%	25

ND = Not detected
NC = Non-calculable
RPD = Relative Percent Difference
SPIKE = Fortified sample

Validation Signature Page

Maxxam Job #: A878412

The analytical data and all QC contained in this report were reviewed and validated by the following individual(s).



TROY CARRIERE, B.Sc., C.Chem, Scientific Specialist

=====

Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.

**DIV METALS
PRESERVATION
CHECKED**

6740 Campobello Rd
Mississauga
Ontario L5N 2L8

17-Jul-08 14:50
JAMES ASPIN
REC
A878412
NNA ENV-998

Tel: 905-817-5779
Fax: 905-817-5779
Toll Free: 1-800-563-6266

**Water Systems are Regulated
Under Reg 243**

SIF : Sample Inspection
Resolved By: _____
Date: _____

Maxxam Analytics Inc.
CAM FCD-00196/2

Report to:

Company Name:	Dufferin-Peel Catholic District School Board		
Client Contact:	Roland Lawton, General Manager of Physical Plant Facilities		
Address:	5685 Keaton Cres., Mississauga ON L5R 3H5		
E-Mail:	<u>roland.lawton@dpcdsb.org</u>	Tel:	905-890-0708 ext. 23204
Project:	Regulation 243	Fax:	905-890-5841
Quote #:	A83456	P.O.#:	BN500636
Lab Service Rep.:	James Aspen		

Invoice to: Same as "Report to"

PAGE _____ OF _____
ANALYSIS REQUESTED

Specify School Type	M.O.E. Assigned Water System Number
School	
Private School	
Day nursery	
	500013820

Sample Identification	Sample Type (Please Specify Standing / Flushed)	Date Sampled	Time Sampled	Sampler Initials	Field Chlorine		No. of Bottles	Lead	School	Private School	Day nursery
					Total	Free					
St. Rose of Lima Kitchen - Staffroom	Standing	July 17/08	6:55 AM				1	X	X		
Kitchen - Staffroom	Flushed	July 17	7:32				1	X	X		

* In accordance with the Safe Drinking Water Act (2002), notification of adverse results must be reported to the SAC, MOH & Waterworks.
Failure to do so is an offence under the act. It is **IMPERATIVE** that the areas shaded in grey be filled out completely before submitting to Maxxam Analytics.

TAT (Turnaround Time)	Notification Information for School, Private Schools & Day Nurseries			LABORATORY USE ONLY	
RUSH TAT MUST HAVE PRIOR APPROVAL	Facility Name: St. Rose of Lima	Medical Officer of Health		Received By: <i>JR. Johnson</i>	
<input checked="" type="checkbox"/> Normal (5-15 Working Days)	Facility Address: 4590 The Gallops	Region: Peel Region		Date: 08/07/08	Time: 14:50
<input type="checkbox"/> RUSH (Specify Below)	Mississauga, ON L5M 2W4	Tel #:		Comments: 12/12/13	
Date Due:	Tel: 905-828-4076	Fax: 905-828-4076	Fax #:	JUL 17 '08 14:50	
PN #	E-mail:	Email:			
Relinquished by (Sign Name): <i>ADAM KNOX</i>	Date: July 17/08	Laboratory Prepared Container:		<input type="checkbox"/> Yes <input type="checkbox"/> No	
Relinquished by (Print Name): ADAM KNOX	Time: 7:32 AM				

NOTE: Samples submitted to Maxxam Analytics must be received below 10°C.
It is the Sampler's responsibility to ensure adequate cooling is provided so that this receiving temperature is met.
Please ensure that a separate COC document is used for each indiv