

Your P.O. #: BN500116
Your Project #: ST SEBASTIAN
Site:500018786
Your C.O.C. #: na

Attention: Roland Lawton
Dufferin Peel Catholic District...
School Board
40 Matheson Blvd W
Mississauga, ON
L5R 1C5

Report Date: 2008/07/28

CERTIFICATE OF ANALYSIS

MAXXAM JOB #: A878353
Received: 2008/07/17, 14:47

Sample Matrix: Water
Samples Received: 2

| Analyses | Quantity | Date Extracted | Date Analyzed | Laboratory Method | Method Reference |
|--|----------|-------------------|------------------|-------------------|---------------------|
| Metals Analysis by ICPMS (as received) 0 | 2 | 2008/07/25 | 2008/07/26 | CAM SOP-00447 | EPA 6020 |

* RPDs calculated using raw data. The rounding of final results may result in the apparent difference.

(1) Metals analysis was performed on the sample 'as received'.

Encryption Key

Please direct all questions regarding this Certificate of Analysis to your Project Manager.

JAMES ASPIN, Project Manager
Email: James.Aspin@maxxamAnalytics.com
Phone# (905) 817-5700 Ext:5771

=====
Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.

For Service Group specific validation please refer to the Validation Signature Page

Total cover pages: 1

Maxxam Job #: A878353
Report Date: 2008/07/28

Dufferin Peel Catholic District...
Client Project #: ST SEBASTIAN
Project name: 500018786
Your P.O. #: BN500116
Sampler Initials: KW

ELEMENTS BY ATOMIC SPECTROSCOPY (WATER)

| | | | | | |
|---------------|--------------|-----------------|--|------------|-----------------|
| Maxxam ID | | | Z96731 | | |
| Sampling Date | | | 2008/07/17 07:05 | | |
| COC Number | | | na | | |
| | Units | Criteria | ST. SEBASTIAN-KITCHEN-STAFFROOM- STANDING | RDL | QC Batch |

| | | | | | |
|---|------|-------------|--------|--------|---------|
| Metals | | | | | |
| . Lead (Pb) | mg/L | 0.01 | 0.0009 | 0.0005 | 1570546 |
| <p>RDL = Reportable Detection Limit QC Batch = Quality Control Batch CRITERIA: Ontario Regulation 243/07 made under the Safe Drinking Water Act, 2002. Released June 7, 2007</p> | | | | | |

| | | | | | |
|---------------|--------------|-----------------|---|------------|-----------------|
| Maxxam ID | | | Z96732 | | |
| Sampling Date | | | 2008/07/17 07:46 | | |
| COC Number | | | na | | |
| | Units | Criteria | ST. SEBASTIAN-KITCHEN-STAFFROOM- FLUSHED | RDL | QC Batch |

| | | | | | |
|---|------|-------------|--------|--------|---------|
| Metals | | | | | |
| . Lead (Pb) | mg/L | 0.01 | 0.0009 | 0.0005 | 1570895 |
| <p>RDL = Reportable Detection Limit QC Batch = Quality Control Batch CRITERIA: Ontario Regulation 243/07 made under the Safe Drinking Water Act, 2002. Released June 7, 2007</p> | | | | | |

Maxxam Job #: A878353
Report Date: 2008/07/28

Dufferin Peel Catholic District...
Client Project #: ST SEBASTIAN
Project name: 500018786
Your P.O. #: BN500116
Sampler Initials: KW

| | |
|-----------|--------|
| Package 1 | 12.3°C |
|-----------|--------|

Each temperature is the average of up to three cooler temperatures taken at receipt

GENERAL COMMENTS

Results relate only to the items tested.

Dufferin Peel Catholic District...
Attention: Roland Lawton
Client Project #: ST SEBASTIAN
P.O. #: BN500116
Project name: 500018786

Quality Assurance Report
Maxxam Job Number: MA878353

| QA/QC Batch | QC Type | Parameter | Date Analyzed yyyy/mm/dd | Value | Recovery | Units | QC Limits |
|-------------|--------------|-------------|-----------------------------|-------------|----------|-------|-----------|
| 1570546 JBW | MATRIX SPIKE | . Lead (Pb) | 2008/07/26 | | 107 | % | 80 - 120 |
| | Spiked Blank | . Lead (Pb) | 2008/07/26 | | 97 | % | 85 - 115 |
| | Method Blank | . Lead (Pb) | 2008/07/26 | ND, RDL=0.5 | | ug/L | |
| | RPD | . Lead (Pb) | 2008/07/26 | NC | | % | 25 |
| 1570895 HRE | MATRIX SPIKE | . Lead (Pb) | 2008/07/26 | | 100 | % | 80 - 120 |
| | Spiked Blank | . Lead (Pb) | 2008/07/26 | | 101 | % | 85 - 115 |
| | Method Blank | . Lead (Pb) | 2008/07/26 | ND, RDL=0.5 | | ug/L | |
| | RPD | . Lead (Pb) | 2008/07/26 | 1.2 | | % | 25 |

ND = Not detected
 NC = Non-calculable
 RPD = Relative Percent Difference
 SPIKE = Fortified sample

Validation Signature Page

Maxxam Job #: A878353

The analytical data and all QC contained in this report were reviewed and validated by the following individual(s).

Cristina Nervo

CHRISTINA NERVO, Scientific Services

=====

Maxxam has procedures in place to guard against improper use of the electronic signature and have the required "signatories", as per section 5.10.2 of ISO/IEC 17025:2005(E), signing the reports. SCC and CAEAL have approved this reporting process and electronic report format.



6740 Campobello Rd
Mississauga
Ontario L5N 2L8

REGULATION 243 DRINKING WATER SUB-CHAIN OF CUSTODY RECORD

Tel: 905-817-5700
Fax: 905-817-5779
Toll Free: 1-800-563-6266

Samples are Regulated Under Reg 243

17-Jul-08 14:47
JAMES ASPIN
A878353
CCO ENV-832

Report to:

Company Name: Dufferin-Peel Catholic District School Board
Client Contact: Roland Lawton, General Manager of Physical Plant Facilities
Address: 5685 Keaton Cres., Mississauga ON L5R 3H5
E-Mail: roland.lawton@dpcdsb.org Tel: 905-890-0708 ext. 23204
Project: Regulation 243 Fax: 905-890-5841
Quote #: A83456 P.O.#: BN500636
Lab Service Rep.: James Aspin

Invoice to: Same as "Report to"

PAGE ____ OF ____

ANALYSIS REQUESTED

Specify School Type

M.O.E. Assigned Water System Number

500018786

| | | |
|-------------------------------------|--------------------------|--------------------------|
| School | Private School | Day nursery |
| <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> |

| Sample Identification | Sample Type (Please Specify Standing / Flushed) | Date Sampled | Time Sampled | Sampler Initials | Field Chlorine | | No. of Bottles | Lead | Specify School Type | | | |
|-----------------------|---|--------------|--------------|------------------|----------------|------|----------------|------|---------------------|----------------|-------------|--|
| | | | | | Total | Free | | | School | Private School | Day nursery | |
| St. Sebastian | | | | | | | | | | | | |
| Kitchen - Staffroom | Standing | Jul 17 | 7:05 | | | | 1 | x | x | | | |
| Kitchen - Staffroom | Flushed | Jul 17 | 7:46 | | | | 1 | x | x | | | |

SIF : Sample Inspection

Resolved By: _____

Date: _____

DW METALS PRESERVATION CHECKED

* In accordance with the Safe Drinking Water Act (2002), notification of adverse results must be reported to the S... Failure to do so is an offence under the act. It is **IMPERATIVE** that the areas shaded in grey be filled out comp...

| | | | | |
|--|--|--|----------------------------------|--|
| TAT (Turnaround Time) | Notification Information for School, Private Schools & Day Nurseries | | LABORATORY USE ONLY | |
| RUSH TAT MUST HAVE PRIOR APPROVAL | Facility Name: St. Sebastian | Medical Officer of Health | Received By: <i>Kim Williams</i> | |
| | Facility Address: 3460 Aquinas Avenue | Region: Peel Region | Date: 08/07/08 Time: 14:47 | |
| <input checked="" type="checkbox"/> Normal (5-15 Working Days) | Mississauga, ON L5M 7L2 | Tel #: | Comments: 12/12/13 | |
| <input type="checkbox"/> RUSH (Specify Below) | | Fax #: | | |
| Date Due: | Tel: 905-607-0107 Fax: 905-607-2535 | E-mail: | | |
| PN # | | | | |
| Relinquished by (Sign Name): <i>Kim Williams</i> | Date: July 17/08 | Laboratory Prepared Container: | | |
| Relinquished by (Print Name): Kim Williams | Time: 3:05pm | <input type="checkbox"/> Yes <input type="checkbox"/> No | | |

NOTE: Samples submitted to Maxxam Analytics must be received below 10°C. JUL 17 '08 14:47
It is the Sampler's responsibility to ensure adequate cooling is provided so that this receiving temperature is met.
Please ensure that a separate COC document is used for each indiv